implox Pty. Ltd.
HEALTHCARE

# CAS 750

# Multi-Parameter Monitor

FOR EVERY LIFE AND BREATH SITUATION™




## User's Manual

CASMED™
FOR WHAT'S VITAL

CE 0086

www.implox.com

## Trademarks

Trademarked names appear throughout this document. Instead of inserting a trademark symbol with each mention of the trademarked name, the publisher states that it is using the names only for editorial purposes and to the benefit of the trademark owner with no intention of improperly using that trademark.

Masimo SET® is a registered trademark of Masimo, Inc. Possession or purchase of this device does not convey any express or implied license to use the device with replacement parts which would, alone, or in combination with this device, fall within the scope of one or more of the patents relating to this device.

Nellcor® and *OxiMax*® are registered trademarks and SatSeconds™ is a trademark of Mallinckrodt, Inc.

Microstream® and FilterLine® are registered trademarks of Oridion Medical 1987 LTD.

YSI® is a registered trademark of Yellow Springs Instrument Company.

MAXNIBP®, Tuff-Cuff®, SoftCheck®, UltraCheck® are registered trademarks and Safe-Cuff™, CASMED™, and *For Every Life and Breath Situation*™ are trademarks of CAS Medical Systems, Inc.

## Contact Addresses

**Corporate Office:**

CAS Medical Systems, Inc.
44 East Industrial Road
Branford, CT 06405
U.S.A.

Phone: +1 (203) 488-6056
in the US: (800) 227-4414

Fax: +1 (203) 488-9438

custsrv@casmed.com
sales@casmed.com
techsrv@casmed.com
www.casmed.com

**Authorized Representative in Europe:**

mossaConsulting GmbH
Bollbrügg 22
23570 Lübeck
Germany

Phone: +49-4502-880-557

Fax: +49-4502-880-559

mossa.rod@t-online.de

Please contact the distributor in the country of purchase if product information or service should be required.

# General Information

This manual is an integral part of the product and describes its intended use. Observance of the manual is a prerequisite for proper product performance and correct operation and ensures patient and operator safety.

The symbol means ATTENTION: Consult accompanying documents.

The warranty does not cover damages resulting from the use of accessories and consumables from other manufacturers.

CAS Medical Systems, Inc. is responsible for the effects on safety, reliability, and performance of the product, only if:

- Assembly, operations, extensions, readjustments, modifications, or repairs are carried out by persons authorized by CAS Medical Systems, Inc.
- The electrical installation of the relevant room complies with the requirements of the appropriate regulations.
- The device is used in accordance with the instructions for use.
- All publications conform to the product specifications and applicable IEC publications on safety and essential performance of electro medical equipment as well as with applicable UL requirements and AHA recommendations valid at the time of printing.

For complete Warranty information, refer to the Warranty Policy located on page 129.

The CAS Medical System, Inc. quality management system complies with the international standards ISO 13485 and the Council Directive on Medical Devices 93/42/EEC.

**Note:** The information in this manual applies only to the CAS 750C and 750E software version 3.0 and the CAS 750S software version 1.0. It does not apply to earlier software versions.

**Note:** Due to continuing product innovation, specifications in this manual are subject to change without notice.

In the U.S. the following caution applies:

**Caution:** Federal law restricts this device to sale by or on the order of a physician or properly licensed practitioner.

**Warning:** Before using the monitor for the first time, please read the information given in section “Safety” starting on page 14.

# About this Manual

**Note:** This manual addresses all parameters a CAS 750 Monitor can have installed. It remains suitable for use if the monitor, that was purchased, has a sub-set of parameters only. Please refer to those sections that are applicable for the model in use.

## Manual Purpose

This manual contains the instructions necessary to operate the CAS 750 monitor safely and in accordance with its functions and intended use.

## Intended Audience

This manual is written for clinical professionals. Clinical professionals are expected to have working knowledge of medical procedures, practices, and terminology as required for monitoring of critically ill patients.

**Caution:** For continued safe use of this equipment, it is necessary that the listed instructions be followed. However, instructions listed in this manual in no way supersede established medical practices concerning patient care.

## Conventions

**Warning:** Directions that warn of conditions that put the patient or the caregiver at risk.

**Caution:** Directions that help to avoid damaging the monitor or losing data.

**Note:** Directions that make it easier to use the monitor, something not readily apparent.

**Figures:** All illustrations in this manual are provided as examples only. They may not necessarily reflect your monitoring setup or data displayed on your monitor.

## Revision History

This manual has a revision number located at the bottom of each page. It changes whenever the manual is updated.
First Printing: 09/2006.

## Read this manual carefully before patient use of the monitor.

CAS Medical Systems, Inc. reserves the right to make changes to this manual and improvements to the product it describes at any time without notice or obligation.

Copyright 2005 CAS Medical Systems, Inc. All rights reserved.
No part of this manual may be reproduced without the written permission of CAS Medical Systems, Inc.

# Contents

Trademarks ....2
Contact Addresses ....2
General Information ....3
About this Manual ....4
Manual Purpose ....4
Intended Audience ....4
Conventions ....4
Revision History ....4

**1 Safety 15**

Indications for Use ....15
Contraindications ....15
Installation and Setup ....16
Device Handling ....18
Safety Checks ....19
Monitoring ....19
Initial Inspection ....20
Monitor Checklist ....21
Patient Environment ....22
Monitor Classifications of Electrical Insulation ....22

**2 Basic Operations 23**

Introduction ....23
Front View ....24
Rear View ....25
Bottom View ....25
Infrared Communication Port ....25
Serial Number Label ....25
Getting Started ....27
Site Requirements ....27
Power Requirements ....27
Turning the Monitor On ....28
Turning the Monitor Off ....29
Power Fail Message ....29
Battery Operation ....29
Battery Status ....30
Battery Conditions ....30

Connecting the Accessories ....31
Front Panel Controls....32
Main Screen ....34
Freeze Traces....36
Pacemaker Indicator ....36
History Screens ....36
Trend History ....37
Alarm History ....38
Printing Trend or Alarm History ....38
Erase Trend History Data....39
Erase Alarm History Data....39
Parameters Menu ....40
$CO_2$ Library ....41
Audio/Visual Menu....41
Auto Dim....42

**3 Alarms 43**

Patient Alarms ....43
Manifestation of Patient Alarms....43
Equipment Alarms ....43
Manifestation of Equipment Alarms....43
Silencing Alarms....44
No alarm active....44
Alarm is active....44
Silence Indication....44
Silence Period is set to “2-Minute” ....44
Silence Period is set to “Permanent” ....44
Alarm Limits....45
Alarm Limits Menu ....45
Changing Alarm Limits....46
Saving Alarm Limits ....46
Restore Alarm Limits....47
Factory Default Alarm Limits ....47
Alarm Volume....48
Alarm Delays ....48
Stand By mode (750S)....49

## 4 ECG/Resp 51

Introduction....51

Safety....51

Preparations....52
- Skin Preparation....52
- Electrode Placement....53
  - 3-Leadwire Electrode Placement....53
  - 5-Leadwire Electrode Placement....54
- Respiration Monitoring....55
  - Electrode Placement for Respiration....55
  - CVA Filter....56
  - Breath-Weighting....56

ECG/Respiration Monitoring....57
- Monitoring Pacemaker Patients....57
- Disconnection of Lead Wires and Patient Cable....58

ECG/Respiration Display Window....59

ECG/Respiration related Settings....59

ECG/Respiration Troubleshooting....60
- ECG/Respiration related Messages....60

## 5 $SpO_2$ 61

Introduction....61

Safety....61

$SpO_2$ Sensors....62

Preparations....63

$SpO_2$ Monitoring....63
- SatSeconds™ Alarm Management....64
- Disconnecting $SpO_2$ Accessories....65

$SpO_2$ Display Window....65

$SpO_2$ related Settings....65

$SpO_2$Troubleshooting....66
- $SpO_2$ related Messages....67

## 6 NIBP 71

Introduction....71

Safety....71

Preparations....73
- Patient Mode Selection....73
- NIBP Hose Selection....73
- Cuff Selection....73
- Cuff Application....74

NIBP Measurements....75
- Starting a Blood Pressure Reading....76
- Stopping a Blood Pressure Reading....76
- Entering the Cycle Mode....77
- Terminating the Cycle Mode....77
- STAT Mode....77
  - Entering STAT Mode....78
  - Exiting STAT Mode....78

NIBP Display Window....79

NIBP Menu....79
- To enter the NIBP Menu....79
- Menu Options....80
- Operating the NIBP Menu....80

NIBP related Settings....80

NIBP Troubleshooting....81
- NIBP related Messages....81

## 7 $CO_2$ 83

Introduction....83

Safety....83

Preparations....84
- Microstream $CO_2$ Consumables....84
- Non-Intubated Application....84
- Intubated Applications....85

$CO_2$ Monitoring....85
- Removing the $CO_2$ Consumables....86

$CO_2$ Display Window....86

$CO_2$ related Settings....87

$CO_2$ Troubleshooting....87
- $CO_2$ related messages....87

**8 TEMP 89**

Introduction....89
Safety....89
Preparations....89
Temperature Monitoring....89
Temperature Display Window....90
Temperature related Settings....90
Temperature Troubleshooting....90
Temperature related Messages....90

**9 Setup 91**

Entering the Setup Menu....91
Selecting the Language....92
Selecting Patient Mode....92
Configure Audio Silencing....93
Setting the Date....93
Setting the Time....94
Daylight Saving Time Option....94
Pacemaker Detection....95
Selecting CVA Filter....95
Selecting Alarm Delay....95
Selecting $CO_2$ Units....96
Selecting Temperature Units....96
Selecting Monitor Set-Up Default....96
Selecting Display Background....97

## 10 Cleaning 99

Cleaning the Monitor ....99
Cleaning Patient Cable and Leadwires ....100
Cleaning Cuffs and Pneumatic Hoses ....100
Cleaning the $SpO_2$ Interconnect Cable ....100
Cleaning $SpO_2$ Sensors ....101
Cleaning Temperature Probes ....101
Cleaning $CO_2$ Consumables ....101
Cleaning the Printer ....101

## 11 Maintenance 103

Maintenance Intervals ....103
Maintenance Checks ....104
- Entering the Service Menu ....104
- Exiting the Service Menu ....104
- IrDA Test ....105
- $CO_2$ Calibration Check ....105
- $CO_2$ Calibration ....106
- NIBP Checks ....108
  - Pneumatic Pressure Check ....108
  - Calibration Check and Overpressure Test ....109
- PIC Voltage ....110

Other Checks ....110
- Temperature Calibration Check ....110

Battery Replacement ....111
- Removing the Battery ....111
- Inserting the Battery ....112

Fuse Replacement ....112
Software Versions ....113
Storage ....114

## 12 Printer 115

Printer Controls and Indicators ....116
Printer Operation ....116
- Direct Connection ....116
- Infrared Connection ....117

Charging the Printer Battery ..... 119
Installing Paper ..... 120
Removing the Battery Pack ..... 121
Installing the new Battery Pack: ..... 122

**13 External Device Interfacing 123**

Nurse Call and RS232 Interface ..... 123
RS232 Interface ..... 123
Nurse Call Interface ..... 123
Mounting ..... 124

**14 Appendix 125**

Symbols ..... 125
Front Panel Symbols ..... 125
Screen Indicators ..... 126
Symbols near Accessory Connections ..... 126
Symbols on Monitor or Printer ..... 127
Symbols on Packaging ..... 128
Warranty Policy ..... 129
Monitor Error Messages ..... 131
Monitor Configurations ..... 132
CAS 750 Models with AC Power Supply ..... 132
CAS 750 Models with 12 VDC Power Input ..... 133
Monitor Configuration Record ..... 134
Specifications ..... 135
Certificates ..... 144
Electronic Emissions and Immunity ..... 144
CE Marking Information ..... 146
Accessories ..... 147
ECG Accessories ..... 147
Tuff-Cuff Blood Pressure Cuffs ..... 147
Safe-Cuff Blood Pressure Cuffs ..... 147
Inflation Hoses ..... 148
Masimo $SpO_2$ Accessories ..... 148
Nellcor $SpO_2$ Accessories ..... 149
Temperature Accessories ..... 149
Capnography Accessories ..... 149
Other Accessories and Options ..... 150

# Figures

Figure 1: Patient Environment ..... 22
Figure 2: Front View ..... 24
Figure 3: Rear View of Monitors ..... 25
Figure 4: Turning the Monitor On ..... 28
Figure 5: Left Side View ..... 31
Figure 6: Front Panel Controls ..... 32
Figure 7: Main Screen ..... 35
Figure 8: Trend History Screen ..... 37
Figure 9: Alarm History Screen ..... 38
Figure 10: Parameters Menu ..... 40
Figure 11: Audio/Visual Menu ..... 41
Figure 12: Alarm Limits Menu ..... 45
Figure 13: Adult and Neonatal Electrode Placement ..... 53
Figure 14: 5-Lead Placement ..... 54
Figure 16: Detaching the Lead Wires ..... 58
Figure 17: Detaching the Patient Cable ..... 59
Figure 18: Cuff Application Range Marker ..... 74
Figure 19: Cuff Application ..... 74
Figure 20: NIBP Menu ..... 79
Figure 21: Setup Menu ..... 91
Figure 22: Service Menu ..... 104
Figure 23: Removing the Battery Pack ..... 111
Figure 24: Software Versions ..... 113
Figure 25: Printer Controls and Indicators ..... 116
Figure 26: History Sample Printouts ..... 118
Figure 27: Waveform Sample Printout ..... 119
Figure 28: Paper Installation ..... 120
Figure 29: Opening the Battery Door ..... 121
Figure 30: Installing the New Battery ..... 122
Figure 31: DB9 Connector ..... 123
Figure 32: Mounting Threads ..... 124

# Tables

Table 1: Factory Default Alarm Limits ..... 47
Table 2: 3-Lead Color and Coding ..... 53
Table 3: 5-Lead Color and Coding ..... 54
Table 4: CAS 750 Models with AC Power ..... 132
Table 5: CAS 750 Models with DC Power ..... 133

This page is intentionally left blank

# 1 Safety

The following Warnings, Cautions and Notes have to be obeyed to guaranty a safe operation of the monitor. Additional Warnings, Cautions and Notes, which apply to specific parameters, are listed in the related sections for each parameter.

## Indications for Use

The 750 Patient Monitor is intended to continuously monitor a patients ECG, heart rate, noninvasive blood pressure (NIBP), functional arterial oxygen saturation ($SpO_2$), respiration rate, temperature and end tidal carbon dioxide ($CO_2$). The monitor is designed as a bedside/portable monitor and is intended for use on adult, pediatric and neonatal patients in the care of health care professionals.

## Contraindications

- ECG electrodes are contraindicated for use on patients with limited skin access or allergic reaction to electrode adhesive or application gel.
- Reusable ECG electrodes are contraindicated for use for prolonged periods of use. It is not intended for long term monitoring. Electrodes must be removed and repositioned if indicated by skin integrity, and reapplied to a different monitoring site.
- Respiration monitoring is contraindicated for patients who are receiving high frequency ventilation assistance.
- Reusable $SpO_2$ sensors are contraindicated for use for prolonged periods of use. It is not intended for long term monitoring. It must be removed and repositioned every four (4) hours and if indicated by circulatory condition or skin integrity, reapplied to a different monitoring site.
- Disposable $SpO_2$ sensors are contraindicated for patients that exhibit allergic reactions to adhesive tape. The sensors must be removed and repositioned every eight (8) hours and if indicated by circulatory condition or skin integrity, reapplied to a different monitoring site.
- No other contraindications are known at this time.

# Installation and Setup

Follow the instructions given in paragraph “Site Requirements” on page 27 and “Power Requirements” on page 27.

**Warning:** The monitor is not intended for diagnostic use. The health care professional should seek a full capability ECG system for diagnostic purposes.

**Warning:** Do not come into contact with patients during defibrillation. Otherwise serious injury or death could result.

**Warning:** The CAS 750 Monitor is intended only as an adjunct in patient assessment. It must be used in conjunction with clinical signs and symptoms.

**Warning:** Do not rely exclusively on the audible alarm system for patient monitoring. Adjustment of alarm volume to a low level during patient monitoring may result in a hazard to the patient. Remember that the most reliable method of patient monitoring combines close personal surveillance with correct operation of monitoring equipment.

**Warning:** Do not use the CAS 750 monitor for any purpose other than specified in this manual. Doing so will invalidate the monitor’s warranty.

**Warning:** Do not connect more than one patient to a monitor. Do not connect more than one monitor to a patient.

**Warning:** *Leakage Current Test* - The interconnection of auxiliary equipment, including a patient monitor or other patient connected equipment, with this device may increase the total leakage current. When interfacing with other equipment, qualified biomedical engineering personnel must perform a test for leakage current before using it with patients. Serious injury or death could result if the leakage current exceeds applicable standards.

**Warning:** Do not use the CAS 750 monitor for Open Heart Applications (Intracardiac Application).

**Warning:** The CAS 750 Monitor is not intended to be used in Oxygen Enriched Atmospheres.

**Warning:** The CAS 750 Monitor is defibrillator proof. It may remain attached to the patient during defib., but the readings may be inaccurate during use and less than ten (10) seconds thereafter.

**Warning:** Do not use the monitor in the presence of Magnetic Resonance Imaging (MRI) equipment.

**Warning:** The CAS 750 Monitor is not “Category AP or APG Equipment”.

**Warning:** Explosion *Hazard* - Do not use the monitor in the presence of a flammable Anesthetic Mixture with Air or with Oxygen or Nitrous Oxide.

**Warning:** *Explosion Hazard* - Do not use this equipment in the presence of flammable anesthetics, vapors or liquids.

**Warning:** *Electromagnetic Compatibility (EMC)* - The equipment needs special precautions if it is placed close to a strong transmitter such as X-ray equipment, MRI devices, TV, AM/FM radios, police/ fire stations, a HAM radio operator, an airport, or cellular phone. Their signals could interfere with the monitor, which may result in disruption of performance of this device or prevents the clear reception of signals by the monitor.

**Caution:** Qualified biomedical engineering personnel only must interface monitoring equipment with other types of medical equipment. Be certain to consult manufacturers’ specifications to maintain safe operation.

**Caution:** Measurements may be affected in the presence of strong electromagnetic sources such as electro surgery equipment.

**Note:** The CAS 750 Monitor is designed for continuous operation.

**Note:** The CAS 750 Monitor is suitable for use in the presence of electro surgery.

**Note:** The CAS 750 Monitor can remain connected to the patient during Cardio Defibrillation. ECG applied parts are “Type CF Defibrillation Proof”. All other applied parts are “Type BF Defibrillation Proof”.

The monitor has been designed to promote patient safety. All equipment parts are protected against the effects of the discharge of a defibrillator. No separate actions are required when using this equipment with a defibrillator.

# Device Handling

**Warning:** To ensure patient safety, do not place the monitor in any position that might cause it to fall on the patient.

**Warning:** Do not lift the monitor by any sensor cable or line as they could disconnect from the monitor, causing the monitor to fall on the patient.

**Warning:** To avoid electric shock or device malfunction, liquids must not be allowed to enter the device. If liquids have entered a device, take it out of service and have it checked by a service technician before it is used again.

**Warning:** The CAS 750 Monitor provides “DRIP-PROOF” level of protection from ingress to moisture.

**Warning:** Do not place liquids on top of the monitor. Do not immerse the monitor or power cord in water or any liquid.

**Warning**: Do not gas sterilize or autoclave the monitor.

**Warning:** Do not touch part of non-medical electrical equipment in the patient environment after removal of covers, connectors etc… without the use of a tool which operate at voltages not exceeding 25 VAC or 60 VDC and the patient at the same time.

**Warning**: Where the integrity of the external protective conductor in the installation or its arrangement is in doubt, EQUIPMENT shall be operated from its INTERNAL ELECTRICAL POWER SOURCE.

**Warning:** Isolation of product from mains can only be achieved by removal of external power cord.

**Warning:** Route all cables away from patient’s throat to avoid possible strangulation.

**Caution:** Pressing the front panel keys with a sharp or pointed instrument may permanently damage the switch membrane. Press the keys using only your finger.

**Caution:** If the monitor is accidentally wetted, take it out of operation. It should be thoroughly dried. To verify the absence of water, a qualified service technician can remove the rear cover.

**Caution:** To avoid the risk of electrical shock, do not remove the back cover. Refer all servicing to qualified personnel.

**Note:** There are no known risks with common disposal of equipment or accessories; however, the disposing of accessories should follow in accordance with local hospital policies. The user should ensure these policies do not conflict with any local, state or federal guidelines.

## Safety Checks

**Warning:** Do not, under any circumstances, perform any testing or maintenance on the monitor or power cord while the unit is being used to monitor a patient. Unplug the power cord before cleaning or servicing the monitor. The operator should not perform any servicing except as specifically stated in this manual.

**Warning:** The functions of the alarm system for monitoring of the patient must be verified at regular intervals.

**Warning:** Periodically, and whenever the integrity of the product is in doubt, test all functions.

**Warning:** Do not use a frayed or damaged power supply cord or any accessory if you notice any sign of damage. Contact CAS Medical Systems for assistance.

**Warning:** The use of Accessory equipment not complying with the equivalent safety requirements of this equipment may lead to a reduced level of safety of the resulting system. Consideration relating to the choice shall include:

- Use of the accessory in the Patient Environment.

- Evidence that the safety certification of the Accessory has been performed in accordance to the appropriate IEC 601-1 and/or IEC 601-1-1 harmonized national standard.

**Caution:** Inspect the monitor, patient cables, sensors and air hose for damage prior to operation. If any damage is noted, the monitor should not be used until it has been serviced. Only personnel authorized to do so by CAS Medical Systems, Inc. should repair the monitor.

**Caution:** If the monitor fails to respond, do not use it until the situation has been corrected by qualified personnel.

## Monitoring

**Warning:** *Conductive Connections* - Extreme care must be exercised when applying medical electrical equipment. Many parts of the human/machine circuit are conductive, such as the patient, connectors, electrodes, and transducers. It is very important that these conductive parts do not come into contact with other grounded, conductive parts when connected to the isolated patient input of the device. Such contact would bridge the patient's isolation and cancel the protection provided by the isolated input. In particular, there must be no contact of the neutral electrode and ground.

**Warning:** If the accuracy of any value displayed on the monitor or printed on a graph strip is questionable, determine the patient's vital signs by alternative means. Verify that all equipment is working correctly.

**Warning:** Only use protected leadwires and patient cables with this monitor. The use of unprotected leadwires and patient cables creates the potential for making an electrical connection to ground or to a high voltage power source which can cause serious injury or death to the patient.

**Caution:** Use only CAS approved accessories and sensors to ensure patient safety and to preserve the integrity, accuracy and the electromagnetic compatibility of the monitor.

**Caution:** Use only CAS approved ECG cables to ensure proper defibrillation protection.

**Caution:** *Electrocautery* - To prevent unwanted skin burns; apply electrocautery electrodes as far as possible from all other electrodes, a distance of at least 15 cm (6 in.) is recommended.

## Initial Inspection

Before unpacking the monitor, inspect the packaging for damage. If there are any signs of damage to the package, a claim should be filed immediately with the shipping agent. It is the receiver's responsibility to notify the carrier's local office to arrange for the pickup of the damaged items. Save the damaged shipping carton as evidence.

Contact your distributor, CAS sales representative, or call CAS Medical Systems, Inc. to report external damage and to arrange for repair or replacement of damaged equipment.

The shipping carton should contain the items listed below. Unpack the monitor and account for each item. Inspect each item for signs of external damage, dents, cracks, scratches, etc. If an item is missing or damaged, contact your distributor, CAS sales representative, or CAS Medical Systems, Inc.

Record the monitor model, serial number and date of purchase at the back of this manual.

# Monitor Checklist

- (1) CAS 750C, E or S Monitor
  Depending on model ordered
- (1) Hospital Grade AC Power Cord or DC Power Cord
  Depending on model ordered
- (1) 3-Lead ECG/Respiration Patient Cable
  For models with ECG installed
- (1) Lead Wire Set
  For models with ECG installed
- (1) Temperature Adapter Cable
  For models with Temperature installed
- (1) $SpO_2$ Interconnect Cable
  For models with $SpO_2$ installed
- (1) $SpO_2$ Finger Sensor
  For models with $SpO_2$ installed
- (2) FilterLine Set, Adult/Pediatric
  For models with $CO_2$ installed
- (2) Smart CapnoLine, Adult
  For models with $CO_2$ installed
- (1) Ten (10) Foot Coiled Inflation Hose
  For models with NIBP installed
- (1) Blood Pressure Cuffs, Adult
  For models with NIBP installed
- (1) Blood Pressure Cuffs, Child
  For models with NIBP installed
- (1) P9 Calibration Kit
  For models with NIBP installed
- (1) CAS 750 Monitor User's Manual

**Caution:** Use only the CAS approved power cord that was shipped with the monitor to preserve the electromagnetic compatibility of the monitor.

**Note:** The monitor is shipped with the appropriate line cord for the country and/or voltage being used.

## Patient Environment

The CAS 750 Monitor has been tested with specific parts of the "system" used within the Patient Environment. These parts are:

- The CAS 750 Monitor
- Appropriate Accessories as listed in section "Accessories" at the back of this manual.
- Line Cord
- Citizen CMP-10 Mobile Printer (purchased from CAS Medical)
- RS232 Interconnect Cable (supplied with printer)
- AC Adapter / Charger, Model TRC-09-1100-M from Group West or equivalent (supplied with printer)




Figure 1: Patient Environment

## Monitor Classifications of Electrical Insulation

The **CAS 750** Monitor (monitor version with integrated AC Power Supply) is a Class I device.

The **CAS 750M** Monitor (monitor version with external DC Power Supply) is a Class II device.

# 2 Basic Operations

## Introduction

The CAS 750 Monitor is a pre-configured monitor that can include the following measurement functions:

- ECG / Respiration / Temperature
- Pulse Oximetry ($SpO_2$)
- Non Invasive Blood Pressure (NIBP)
- Capnography ($CO_2$)

The ECG signal and the thorax impedance are measured through the same set of electrodes. Patients can be monitored with a three-lead or five-lead ECG cable. ECG and respiration signals can be displayed as waveforms. Temperature is obtained using a thermistor applied to the patient's skin.

The Pulse Oximeter function continuously monitors and displays values for functional arterial hemoglobin saturation (%$SpO_2$) and the pulse rate (PR). The dynamic pulse signal can be displayed as a waveform (Plethysmogram) or as a bar graph indicator.

The non-invasive blood pressure measurement uses the oscillometric method and applies for neonatal, pediatric and adult patients. The user can select between manual, cycle or continuous (STAT) mode.

The MicroStream Capnography provides continuous monitoring of the $EtCO_2$ value and the Respiration Rate (RR) of intubated and non-intubated patients. The $CO_2$ signal can be displayed as a waveform.

The monitor is equipped with a rechargeable battery pack and can be used independently from an external power source.

## Front View




Figure 2: Front View

## Rear View




Figure 3: Rear View of Monitors

(M-Version shown below)

1. Battery Compartment
2. External Device Interface (RS232)
3. Receptacle for the AC power cord
4. Fuse Compartment
5. Receptacle for the 12 VDC cable

## Bottom View

### Infrared Communication Port

An Infrared (Ir) output port, located on the bottom panel of the monitor's front cover, is available to print Waveforms, History and Alarm data to the optional external printer or other data collection device(s). Maximum distance is approximately 3 feet, direct line of sight operation.

### Serial Number Label

The serial number label is located on the bottom of the monitor.

This page is intentionally left blank

# Getting Started

## Site Requirements

**Note:** Put the monitor in a location where you can *easily* see the screen and access the operating controls.

**Caution:** *Stacking* - Where monitor is used adjacent to or stacked with other equipment, the monitor should be observed to verify normal operation in the configuration in which it will be used.

**Caution:** *Negligence* - CAS Medical Systems Inc. does not assume responsibility for damage to the equipment caused by improperly vented cabinets, improper or faulty power, or insufficient wall strength to support equipment mounted on such walls.

## Power Requirements

The external power for the CAS 750 Monitor can be either AC mains power or DC power. The two different models are:

- Model **CAS 750M** for +12V VDC
- Model **CAS 750** for 110 to 240 VAC

The following applies for the CAS 750 model with AC power supply:

**Warning**: Do not plug the monitor into an outlet controlled by a wall switch or dimmer.

**Warning**: Where the integrity of the external protective conductor in the installation or its arrangement is in doubt, EQUIPMENT shall be operated from its INTERNAL ELECTRICAL POWER SOURCE.

**Warning:** Isolation of product from mains can only be achieved by removal of external power cord.

**Caution:** Do not defeat the three-wire grounding feature of the power cord by means of adaptors, plug modifications, or other methods. Do not use extension cords of any type.

**Note:** The monitor is suitable to be connected to public AC mains power.

### Turning the Monitor On

Press the POWER key on the front panel to turn the monitor on.




Figure 4: Turning the Monitor On

Each time the monitor is turned on an internal self-test is conducted to ensure that all circuits are functioning properly. Verify the following:

- The monitor produces 3 one-second audio beeps.
- Both, the Equipment and the Alarm Indicators are lit for one-half second and cycled one after the other.
- The monitor displays temporarily its current configuration data:
    - Model number (750C, E or S)
    - Power source (E=External AC, EM=External DC)
    - Number of installed parameters
    - Brand of installed parameters (MS = Masimo, NL = Nellcor, C = Oridion)

**Note:** If after the Power-On-Self-Test is completed the monitor displays “NIBP Cal” in the Equipment Message area, the monitor should be returned to CAS Medical Systems for service.

Once the Power-On-Self-Test is completed, the monitor’s Main screen is activated and the monitor is ready for use.

## Turning the Monitor Off

When the monitor is not being used, it may be turned off by pressing the POWER key for 2 seconds. The display will turn blank and the unit is no longer monitoring the patient.

**Note:** The internal power supply is not switched off! It remains connected to the external power source (AC or DC) to enable battery charging.

### Power Fail Message

During the internal self-test the monitor checks whether it was previously turned off correctly or if it was disconnected from power in any other way. The user will be alerted about such a power loss after the monitor is turned on again.

The message “Power Failure” is displayed in the Equipment Message Window, the Equipment Alarm LEDs are activated and 3 beeps followed by 2 beeps are heard every 10 seconds.

- Press the SILENCE/RESET key to clear the Power Fail condition.

**Note:** Only the SILENCE/RESET and the ON/STANDBY keys are operable, all other keys are disabled while the message is displayed.

## Battery Operation

The monitor is equipped with a rechargeable battery. The battery is charging whenever the monitor is connected to an external power source (AC Line Power or +12 VDC).

Batteries will self-discharge when they are not used. It is recommended leaving the monitor connected to an external power source whenever possible.

The monitor will operate on a completely charged battery for 3 to 5 hours depending on its configuration and the use of the NIBP function.

### Battery Status

There are several means to get an indication about the battery charge status:

- When the monitor is connected to external power and turned off, the Main screen displays the Battery Indicator icon with a moving bar from left to right within the indicator signifying the battery is being charged. Once charged, the moving bar will stop and the battery icon will be completely filled in.
- When the monitor is being powered from the battery, the moving bar within the Battery Indicator icon will be moving from right to left signifying the battery is being discharged.

### Battery Conditions

The user will be alerted in case the battery charge level gets low:

- The "Low Battery" or "Dead Battery" message is displayed.
- The Power Indicator changes from orange to red.
- The Equipment Alarm LEDs are activated.
- An audible tone is generated.

**Warning:** When the "Dead Battery" message is displayed, the patient is no longer being monitored. After approx. 3 minutes in the "Dead Battery" condition, the monitor will turn itself off.

**Warning:** If the battery is not charged, the monitor may no longer function as intended.

**Note:** When the "Low Battery" or "Dead Battery" message appears, the monitor should be connected to an external power source. A depleted battery may be fully recharged in 5 hours.

**Note:** The monitor can remain in normal operation while the battery is charging. During charging, the case may feel warm to the touch.

**Caution:** Under various state and local laws, it may be illegal to dispose of the battery into the municipal waste stream. Check with your local authorities for instructions on recycling options in your area.

# Connecting the Accessories

Connect the various accessories to the appropriate input connector at the left side of the monitor.




Figure 5: Left Side View

1. ECG/Respiration Input Connector
2. $SpO_2$ Probe Connector
3. NIBP Hose Connection
4. $CO_2$ Scavenger Exhaust
5. Temperature Probe Input
6. MicroStream™ $CO_2$ Input Connector

# Front Panel Controls




Figure 6: Front Panel Controls



Power Indicator.

*Green:* Monitor operates on external power.

*Yellow:* Monitor operates on battery power.

*Red:* Low or dead battery.

POWER

*On:* Turns on the monitor's display.

*Standby:* Switches monitor to standby mode when pressed for 2 seconds.

ALARMS

*Reset:* When pressed once during an active patient alarm, silences the audio portion of that alarm for fifteen (15) seconds.

When pressed during an active equipment alarm, the alarm condition shall be acknowledged along with the audio and visual shall be removed.

*Silence*: Starts a period of silence when pressed while no alarms are active. The duration of the silence period depends on its previously made configuration.

**Note:** Press the key twice to enter a silence period while an alarm is active.

NIBP

*Start*: Initiates a NIBP measurement.

*Cancel*: Terminates any active NIBP measurement and immediately deflates the cuff.

*Menu:* Activates the NIBP menu when pressed and held for 2 seconds.



RETURN

*Main Screen*: Returns to the Main screen when pressed while any other screen or menu is being displayed.

*Freeze*: Freezes all traces when pressed while the Main screen is active. Press again to un-freeze. Traces will un-freeze automatically after 60 seconds.

*Print*: Activates a print of the actual screen content (Traces, Trend or Alarm History) when key is pressed for 2 seconds.

Refer to paragraph “Main Screen” on page 34 for more information.

PARAMETERS

Activates the Parameters menu, which provides an overview of parameters and screen layout settings and gives access to change these settings. Refer to paragraph “Parameters Menu” on page 40 for more information.



LIMITS

Activates the Alarm Limits menu, which provides an overview of all actual limit settings and gives access to change, save and restore alarm limits. Refer to section “Alarms” starting on page 43 for more information.

HISTORY

Activates the Trend History and the Alarm History screen. Gives access to erase History data. Refer to paragraph “History Screens” on page 36 for more information.

AUDIO/VISUAL

Activates the Audio/Visual menu which allows the user to configure the audio and visual signals the monitor can generate. Refer to paragraph “Audio/Visual Menu” on page 41 for more information.

UP

Moves cursor upward or scrolls through menu options, press and hold for quicker advance.

Sets patient mode to Adult when pressed and hold while the monitor is being turned on.

DOWN

Moves cursor downward or scrolls through menu options, press and hold for quicker advance.

Sets patient mode to Neonate when pressed and hold while the monitor is being turned on.

### **NEXT** Function

In the menus: The HISTORY and AUDIO/VISUAL keys are programmed to advance horizontally to the ***next*** parameter selection.

### **PREVIOUS** Function

In the menus: The PARAMETERS and LIMITS keys are programmed to move backwards horizontally to the ***previous*** parameter selection.

## Main Screen

**Note:** When switching from the Main screen to any other screen or menu, the monitor will continue to update and display the numeric values of the parameters being monitored.

**Note:** The actual displayed information depends on the parameter configuration of the monitor and the user defined screen layout.




Figure 7: Main Screen

1. Equipment Alarm Indicator
2. ECG Gain
3. Main Display Screen
4. Value for ECG Cal Marker
5. Waveform Selections
6. Cal Marker for ECG and Respiration
7. Time (24 hour format)
8. HR High or Low Alarm Limit Off indicator
9. Battery Indicator
10. Patient Alarm Indicator
11. Silence Indicator

    The LED lighting scheme indicates the selected functionality:

    *Continuous*: 2-Minute audio silence.

    *Flashing, 1s*: Permanent audio silence.

12. Bar Graph Indicator for signal strength when trace not displayed.
13. Operating Mode / Equipment Message Window
14. Respiration Breath Icon, flashes for every detected breath.
15. Patient Message Window
16. NIBP Mode (Manual, Initial Inflation Pressure or Automatic)
17. Heart Beat Icon, flashes for every detected heartbeat.

### Freeze Traces

While the Main screen is being displayed, the user can freeze the traces.

- Press the RETURN key.

The message “Traces Frozen” appears at the top of the Main screen. While the traces are frozen, the numerics continue to update. Traces will automatically un-freeze if no key is pressed for 60 seconds or in case any other screen or menu is entered.

- Press the RETURN key again to manually un-freeze traces.

### Pacemaker Indicator

When the Pacemaker Detection is enabled and a pacemaker impulse is detected, an artificial spike is added to the ECG waveform and the letter “P” is displayed right next to the heartbeat icon. For more information, refer to paragraph ”Monitoring Pacemaker Patients” on page 57.

## History Screens

The monitor collects History data over a 24-hour period. Continuously measured parameters such as HR, RR or %$SpO_2$, are stored as one-minute averages. NIBP readings and alarm events are stored as they occur.

**Note:** Turning the power off does not clear the stored data. The stored data will remain in memory for 24 hours. Older data is deleted automatically. It is suggested to manually clear History Data between patients. Refer to paragraph “Erase Trend History Data“ on page 39.

**Note:** The monitor uses an internal Real Time Clock to time stamp all entries. Changes made to either the time or date settings, should be performed in-between patients being monitored.

**Note:** The information being displayed depends on the monitor configuration and patient specific data.

History data is presented in two screens, the Trend History and Alarm History screen.

- Once the History screen has been displayed, press and hold the HISTORY key for 2 seconds to toggle between the Trend History and the Alarm History screens.
- Press the RETURN key to exit from the History screen.

**Note:** If no key is pressed for 30 seconds, the monitor will automatically exit to the Main screen.

## Trend History

The Trend History screen shows patient data in a tabular form. All continuously monitored parameters are listed as 1-minutes average values. Each line represents one minute. NIBP parameters are displayed in an additional row positioned at the time of occurrence.

- Press the HISTORY key to enter this screen.

**Note:** If there is no data available, the message “No History” is displayed at the top of the screen.

History

| HR:MN | HR | RR | %O2 | PR | CO2 | RR |
|---|---|---|---|---|---|---|
| 13:40 | 60 | 15 | 98 | 59 | 38 | 16 |
| 13:39 | NIBP=120/ 80( 90) | | | PR= 59 | | |
| 13:39* | 60 | 15 | 98 | 59 | 38 | 16 |
| 13:38 | 60 | 15 | 98 | 59 | 38 | 16 |
| 13:37 | NIBP=112/ 78( 89) | | | PR= 61 | | |
| 13:37 | 60 | 15 | 98 | 59 | 38 | 16 |
| 13:36 | 60 | 15 | 98 | 59 | 38 | 16 |
| 13:35 | NIBP=120/ 80( 90) | | | PR= 58 | | |
| 13:35 | 60 | 15 | 98 | 59 | 38 | 16 |

Erase All History No

Figure 8: Trend History Screen

The cursor position is automatically set to the first line of data, which is the most recent data.

- Press the DOWN key to scroll down (back in time)
- Press the UP key to scroll up (forward time).

## Alarm History

The Alarm History screen shows patient alarms in a tabular form as they occurred. Each line represents one alarm event.

- Press the HISTORY key to access the Trend History screen first.
- Then press and hold the HISTORY key for 2 seconds to enter the Alarm History screen.

**Note:** If there is no data available, the message “No Alarm History” is displayed at the top of the screen.




Figure 9: Alarm History Screen

The cursor position is automatically set to the first alarm, which represents the most recent event.

- Press the DOWN key to scroll down (back in time)
- Press the UP key to scroll up (forward in time).

## Printing Trend or Alarm History

To print the Trend or Alarm History data, proceed as follows:

- Enter the Trend or Alarm History screen by pressing the HISTORY key.
- Press and hold the RETURN key for 2 seconds. The message “Printing” appears in the Equipment Message Window.

**Note:** Refer to section “Printer” for more information about the printer and a sample printout of the History screen.

## Erase Trend History Data

**Note:** In-between patients, the previously stored data should be erased to avoid any potential misinterpretation of data.

**Note:** Trend and Alarm History data are erased at the same time.

Follow these steps to erase History data:

- Press the HISTORY key to access the Trend History screen.
- Press the HISTORY key to move the cursor to the line labeled Erase All History No.
- Use the UP and DOWN key to scroll through the available options. Select “Erase All History Yes” and press the HISTORY key.

The monitor erases all History data in memory, exits the History menu and returns to the Main screen.

**Note:** When erasing Trend History, the Initial Inflation pressure is reset to its default value based on Patient type, and the appropriate Alarm Limit set is restored.

**Note:** Turning the monitor off or disconnecting it from all power sources does not erase the History data.

## Erase Alarm History Data

**Note:** In-between patients, the previously stored data should be erased to avoid any potential misinterpretation of data.

**Note:** Alarm History data can be erased separately.

Follow these steps to erase Alarm History data:

- Press the HISTORY key to access the Trend History screen.
- Then press and hold the HISTORY key for 2 seconds to enter the Alarm History screen.
- Press the HISTORY key to move the cursor to the line labeled Erase Alarm History No.
- Use the UP and DOWN key to scroll through the available options. Select “Erase Alarm History Yes” and press the HISTORY key.

The monitor erases all data in Alarm History and returns to the Trend History screen.

**Note:** Turning the monitor off or disconnecting it from all power sources does not erase the History data.

# Parameters Menu

The Parameters menu shows the settings of all parameters and allows changing them.

- Press the PARAMETERS key to enter this menu.

**Note:** The information being displayed depends on the monitor configuration and patient specific settings.

**Note:** The monitor will automatically return to the Main screen if no key is pressed for 30 seconds.




Figure 10: Parameters Menu

The cursor position is automatically at the first item in the left column.

- Use the UP and DOWN keys to select the item that needs to be changed.
- Use the NEXT key to move the cursor to the next column to the right.
- Use the UP and DOWN key to scroll through the available options.

**Note:** Several changes may be done in one session. Use the PREVIOUS key to return the cursor back to the left column and repeat the steps as before.

- Press the RETURN key when finished. All changes will be saved and become immediately effective.

**Note:** A continuously updating Trend History or a menu of NIBP event readings can be chosen, in the Parameters Setup menu as Trace 3.

### $CO_2$ Library

If $CO_2$ is selected as Trace 2 or 3, a library of 10 $CO_2$ traces can be selected for educational viewing. Use the UP key to enter the library, and then select other traces by using the UP or DOWN keys. Traces are displayed at 3 mm/Sec. The $CO_2$ Library will be displayed for 120 seconds and then automatically return to the Main screen.

## Audio/Visual Menu

The Audio/Visual menu shows the actual settings and allows changing them.

- Press the AUDIO/VISUAL key to enter this menu.

**Note:** The monitor will automatically return to the Main screen if no key is pressed for 30 seconds.




Figure 11: Audio/Visual Menu

The cursor position is automatically at the first item in the left column.

- Use the UP and DOWN keys to select the item that needs to be changed.
- Use the NEXT key to move the cursor to the next column to the right.
- Use the UP and DOWN key to scroll through the available options.

**Note:** Several changes may be done in one session. Use the PREVIOUS key to return the cursor back to the left column and repeat the steps as before.

- Press the RETURN key when finished. All changes will be saved and become immediately effective.

## Auto Dim

The CAS 750 Monitor incorporates an Auto Dim feature that, when enable, turns the display's backlight automatically off after 30 seconds. This feature can be helpful when the monitor is being used at night in a patient's room. The Auto Dim feature is enabled when the user selects "ON" in the Auto Dim parameter selection of the Audio/Visual menu.

When the Auto Dim feature is enabled and the LCD screen backlight has turned off, the monitor will flash the top left yellow LED of the Equipment Alarm Light Bar in sync with either the patient's ECG heart rate or $SpO_2$ pulse rate. (Based on the selection made for Beep Source in the Audio/Visual menu).

# 3 Alarms

The monitor distinguishes between patient and equipment alarms.

## Patient Alarms

**Warning**: Before each use, verify that the alarm limits are appropriate for the patient being monitored.

**Note:** All patient alarms, based on **continuously** monitored parameters (e.g. Heart Rate, $SpO_2$, etc.) will clear automatically when the alarm cause is no longer persistent.

**Note:** All patient alarms, based on **discontinuously** monitored parameters (e.g. NIBP readings) will clear only after acknowledged by the user.

### Manifestation of Patient Alarms

- Numeric display of the alarming parameter flashes.
- Patient Alarm LED's are activated.
- Audible tone is generated (3 beeps followed by 2 beeps every 10 seconds).
- Flashing text in Patient Message Window explains the cause.

**Note:** When an ongoing patient alarm is acknowledged by pressing the ALARMS key, only the audible signal will be silenced for 15 seconds. All other alarm indicators will remain active until the alarm condition no longer exists.

## Equipment Alarms

**Note:** All equipment alarms will clear automatically when the alarm cause is no longer persistent.

### Manifestation of Equipment Alarms

- Alarming parameter display flashes.
- Equipment Alarm LED's are activated.
- Audible tone is generated (3 beeps every 25 seconds)
- Flashing text in Equipment Message Window explains the cause.

**Note:** When an ongoing equipment alarm is acknowledged by pressing the ALARMS key, only the audible signal will be disabled. All other alarm indicators will remain active until the alarm condition no longer exists.

**Note:** Low Battery and Dead Battery alarms cannot be silenced.

# Silencing Alarms

In case the monitor should temporarily not generate any alarm tone, the user can activate a Silence Period.

**Note:** When a Silence Period is being entered, the monitor will not generate an audible signal for patient or equipment alarms with the exception of “Low Battery” and “Dead Battery” conditions.

### No alarm active

Pressing the ALARMS key will immediately activate a Silence Period.

### Alarm is active

Pressing the ALARMS key will only acknowledge the active alarm and disable the audio signal for 15 seconds. A second keystroke is required to activate a Silence Period.

### Silence Indication

The duration of the Silence Period can be selected in the monitors Configuration menu. Refer to paragraph “Configure Audio Silencing” on page 93 for more information. Depending on the selection made, the indication for a Silence Period will be as following:

**Silence Period is set to “2-Minute”**

- The Silence Indicator will illuminate continuously.
- The message “2 Minute” will be displayed in the Equipment Message Window.
- The Silence Period will be terminated automatically after 2 minutes.

**Silence Period is set to “Permanent”**

- The Silence Indicator will flash (1 second on/ 1 second off).
- The message “Permanent” will be displayed in the Equipment Message Window.
- To re-activate the audio alarm, the ALARMS key has to be pressed manually.

# Alarm Limits

The monitor can store two independent sets of alarm limit values for both Adult and Neonate modes. These sets and the factory defaults can be restored when needed.

**Note:** Switching between Adult and Neonate mode automatically activates the most recently stored set for patient alarm limits of the appropriate type.

## Alarm Limits Menu

The Alarm Limits menu shows the actual settings of all alarm limits and allows changing them.

- Press the LIMITS key to enter this screen.

**Note:** The information being displayed depends on the monitor configuration and patient specific data.

**Note:** The monitor will automatically return to the Main screen if no key is pressed for 30 seconds.

| Adult 1 | Limits | | |
|---|---|---|---|
| | Low | High | |
| HR | OFF | 220 | BPM |
| %$SpO_2$ | 88 | OFF | % |
| SatSeconds(1) | OFF | | |
| $EtCO_2$ | OFF | OFF | mmHg |
| $FiCO_2$ | | 7 | mmHg |
| RR | OFF | OFF | BrPM |
| No RESP | 30 | | Sec |
| SYS | OFF | 240 | mmHg |
| DIA | OFF | 130 | mmHg |
| Save | No | | |
| Restore | No | | |

Figure 12: Alarm Limits Menu

**Note:** Depending on the monitor configuration, a scroll bar will be displayed indicating that more selections are available to view.

(1) SatSeconds Limit selection only available with Nellcor $SpO_2$ oximetry option.

## Changing Alarm Limits

The cursor is automatically positioned at the first item in the left column.

- Use the UP and DOWN keys to select the alarm limit that needs to be changed.
- Use the NEXT key to scroll to the next “Low” column or advance to the “High” column by pressing the key again.
- Use the UP or DOWN keys to increase or decrease the alarm limit value.

**Note:** Several changes may be done in one session. Use the PREVIOUS key to move the cursor back to the left column and repeat the steps as before.

- Press the RETURN key when finished.

The monitor returns to the Main screen, all changes will become effective immediately and remain valid until the monitor is turned off.

## Saving Alarm Limits

To save the alarm limit changes, proceed as follows:

- Use the DOWN key and move the cursor to the “Save” position.
- Use the NEXT key and move the cursor to the “No” position.
- Use the UP or DOWN keys and select the appropriate alarm limit set “Adult 1” or “Adult 2” for saving.

**Note:** Saving options will be “Neo 1” and “Neo 2” when the monitor is in Neonate mode.

- Press the LIMITS key to start the saving process.

The message “Saving” appears and the alarm limit values will be stored in memory.

If a second set of alarm limits is required, repeat these steps and save the changes under a different name (e.g. “Adult 2” or “Neo 2”).

- Press the RETURN key when finished.

The monitor returns to the Main screen, all changes will become effective immediately and remain valid until alarm limits changed manually or a different set of alarm limits is selected.

## Restore Alarm Limits

To restore a set of alarm limits, proceed as follows:

- Use the DOWN key and move the cursor to the “Restore” position.
- Use the NEXT key and move the cursor to the “No” position.
- Use the UP or DOWN keys and select the appropriate alarm limit set “Factory Default”, “Adult 1” or “Adult 2” to restore.

**Note:** Restore options will be “Factory Default”, “Neo 1” and “Neo 2” when the monitor is in Neonate mode.

- Press the LIMITS key to start the restore process.

The message “Restoring” appears and the alarm limit values are overwritten with the restored values.

- Press the RETURN key when finished.

The monitor returns to the Main screen and the new limit set will become effective immediately.

## Factory Default Alarm Limits

Factory Default Alarm Limits are different for Adult and Neonatal patients.

| | Adult | | Neonate | |
|---|---|---|---|---|
| **Parameter** | **Low** | **High** | **Low** | **High** |
| Heart Rate | OFF | 220 | OFF | 220 |
| %$SpO_2$ | 88 | OFF | 88 | 96 |
| Pulse Rate | OFF | 220 | OFF | OFF |
| SatSeconds | OFF | N/A | OFF | N/A |
| $EtCO_2$ | OFF | OFF | OFF | OFF |
| $FiCO_2$ | N/A | 7 | N/A | 5 |
| Resp Rate | OFF | OFF | OFF | OFF |
| No Resp | 30 | N/A | 20 | N/A |
| Systolic | OFF | 240 | OFF | 120 |
| Diastolic | OFF | 130 | OFF | 80 |

Table 1: Factory Default Alarm Limits

## Alarm Volume

The alarm tone volume can be changed but not completely turned off.

- To change the alarm tone volume follow the steps described in paragraph “Audio/Visual Menu” on page 41.

## Alarm Delays

To reduce the number of false-positive alarms, some patient alarms will be activated after a short delay. These are:

- Heart Rate alarms: 5 seconds
- $EtCO_2$ alarms: 10 seconds
- $FiCO_2$ alarm: 10 seconds
- RR alarms: 10 seconds

When the monitor incorporates *Nellcor* $SpO_2$:

- $\%SpO_2$ alarms: 0 or SatSeconds (configurable)

To change the alarm delay, follow the steps described in paragraph “Changing Alarm Limits” on page 46.

- $SpO_2$ Pulse Rate alarms: 0 or 10 seconds (configurable)

To change the alarm delay, follow the steps described in paragraph “Selecting Alarm Delay” on page 95.

When the monitor incorporates *Masimo* $SpO_2$:

- $\%SpO_2$ and Pulse Rate alarms: 0 or 10 seconds (configurable)

To change the alarm delay, follow the steps described in paragraph “Selecting Alarm Delay” on page 95.

**Note:** All other alarms are generated without any delay.

### Stand By mode (750S)

This mode can be used just prior to connecting the monitor to a patient, but all the monitor's cables are connected.

Press and holding for two (2) seconds the SILENCE/RESET pushbutton.

When enabled, all Equipment alarms are silenced and the message "Stand By" is displayed on the Main display.

For the Stand by feature to work properly, the following conditions must apply;

The monitor cannot be actively monitoring any patient parameter.

To return the monitor to normal operation, connect the $SpO_2$ finger probe to the patient or press the NIBP pushbutton key.

This page is intentionally left blank

# 4 ECG/Resp

## ECG and Respiration

## Introduction

This section describes the ECG and Respiratory monitor function. Both, the ECG signal and the respiratory signal (based on the thorax impedance change) are measured through the same set of ECG electrodes.

The monitor automatically detects whether a 3-lead or a 5-lead patient cable is connected and allows the user to make ECG lead selections accordingly.

The CAS 750 Monitor determines respiration by impedance pneumography. Patient respiration is achieved by applying a low voltage, high frequency AC signal across the active ECG leads (LA/L and RA/R). The monitor detects changes in thoracic impedance that occur as a result in chest movements. Impedance normally increases with inspiration and decreases with expiration. Respiration detection is based on an inspiration *and* an expiration. After an inspiration is detected, an expiration must be detected within 3 seconds (Neonate mode) or 5 seconds (Adult mode) to be counted as a breath.

**Note:** Special considerations are required for patients with pacemaker. Refer to paragraph “Monitoring Pacemaker Patients” on page 57.

## Safety

**Warning:** If uncertain about the accuracy of any measurement, check the patient’s vital signs by alternate means and then make sure the monitor is functioning correctly.

**Warning:** Even though the ECG patient circuit is electrically isolated, it has not been designed for direct application on a patient’s heart.

**Warning:** The monitor is not detecting arrhythmias and is not alarming on irregular ECG rhythms.

**Caution:** ECG electrodes are contraindicated for use on patients with limited skin access or allergic reaction to electrode adhesive or application gel.

**Caution:** Reusable ECG electrodes are contraindicated for use for prolonged periods of use. It is not intended for long term monitoring. Electrodes must be removed and repositioned if indicated by skin integrity, and reapplied to a different monitoring site.

**Caution:** Do not rely on the ECG waveforms for any diagnostic purposes.

# Preparations

**Caution:** Prior to patient monitoring, ensure the monitor is configured to the appropriate patient mode – Neonate or Adult. Refer to paragraph "Selecting Patient Mode" on page 92.

**Note:** The quality of ECG information is a direct result of the quality of the electrical signal received at the electrode.

**Note:** Proper skin preparation is necessary for good signal quality at the electrode. A good signal at the electrode provides the monitor with valid information for processing the ECG data.

**Note:** Always check the Date Code of the electrodes prior to applying them to the patient.

**Note:** Do not use the 5-Lead Patient Cable for 3-Lead monitoring. A "Leads OFF" message would be displayed.

## Skin Preparation

The following is a suggested guideline for skin preparation and should be followed for all electrode types.

- Choose flat, non-muscular areas to place electrodes.
- Make sure the skin area where the electrodes are to be placed is clean, dry, intact and free of powder, oil or lotion.
- If necessary, shave hair from skin at chosen sites.
- Using a dry washcloth, gently rub skin surface at sites to remove dead skin cells.
- Using a washcloth, thoroughly cleanse the site with a mild soap and water solution. Be sure to remove all oily residue, dead skin cells, and abrasives. Leftover abrasion particles can be a source of noise (artifact).

## Electrode Placement

Special consideration may be required if the ECG electrodes are used to detect respiration. Follow the guidelines given in paragraph “Electrode Placement for Respiration” on page 55.

**Note:** When using “snap” leadwires, attach leadwires to electrodes first then apply electrodes to the patient. This prevents the gel from spreading and becoming ineffective as you attach the snaps to the electrodes.

### 3-Leadwire Electrode Placement

The lead wires that attach to the electrodes are color coded for ease of identification.

| 3-Lead AAMI Standard | 3-Lead IEC Standard |
|---|---|
| LA = black (left arm) | L = yellow (left arm) |
| RA = white (right arm) | R = red (right arm) |
| LL = red (left leg) | F = green (foot) |

Table 2: 3-Lead Color and Coding

The two active electrodes (left and right arm) should be placed on the patient’s thorax as shown in the following picture. The third electrode (left leg) is a reference electrode allowing for better signal detection.




Figure 13: Adult and Neonatal Electrode Placement

### 5-Leadwire Electrode Placement

The lead wires that attach to the electrodes are color coded for ease of identification.

| 5-Lead AAMI Standard | 5-Lead IEC Standard |
|---|---|
| LA = black (left arm) | L = yellow (left arm) |
| RA = white (right arm) | R = red (right arm) |
| RL = green (right leg) | N = black (neutral) |
| LL = red (left leg) | F = green (foot) |
| V = brown (common) | C = white (chest) |

Table 3: 5-Lead Color and Coding




Figure 14: 5-Lead Placement

The placement of the common V/C electrode is described in the following:

1) Fourth intercostals space to the right of the sternum
2) Fourth intercostals space to the left of the sternum
3) Directly between leads 2 and 4
4) Fifth intercostals space at midclavicular line
5) Level with 4 at left anterior auxiliary line
6) Level with 5 at left midaxillary line

## Respiration Monitoring

**Warning:** Impedance respiration monitoring is not reliable when ECG electrodes are placed on the limbs.

**Warning:** The monitor may not detect all episodes of inadequate breathing, nor does it distinguish between central, obstructive and mixed apnea events.

**Caution:** Respiration monitoring is contraindicated for patients who are receiving high frequency ventilation assistance.

Several considerations have to be made when the patient's respiration is monitored.

### Electrode Placement for Respiration

Electrode placement is crucial to monitoring respiration by the impedance method. The sensitivity of the monitor and its ability to accurately detect respiration is greatly enhanced or impeded by the quality of the electrodes and optimal electrode placement.

**Note:** The respiration signal is received from the LA/L and RA/R electrodes. These leads are available on 3-lead and 5-lead cables.

- Observe the patient and place the electrodes where the greatest breathing movement occurs on the chest.

**Note:** Should Respiration monitoring be difficult due to lead placement or too much motion artifacts, the Respiration function can be turned off. Refer to paragraph "Parameters Menu" on page 40.

**Note:** If $CO_2$ monitoring is active, the Respiration Rate Alarm value will be determined from the $CO_2$ measurement.

### CVA Filter

Since impedance monitors are very sensitive to chest wall movements, the possibility of mistaking changes due to cardiac activity exists. A Cardiovascular Artifact (CVA) filter prevents apnea episodes from being missed due to cardiogenic artifact. The CVA filter simultaneously monitors the frequency of both the cardiac and respiratory signals. When enabled any deflections in the respiratory waveform that occur at the same frequency as the ECG or higher are filtered out before the monitor sends the signal to the breath detector. Refer to paragraph “Selecting CVA Filter” on page 95.

### Breath-Weighting

**Note:** The Breath-Weighting functionality is automatically active when the monitor is in Neonate mode, it cannot be disabled.

**Note:** Breath-Weighting is not applied when the monitor is in Adult mode.

A “No RESP” alarm is generated if the next breath does not occur before a specified time limit (e.g. 30 seconds) is exceeded. The monitor uses an internal timer to accomplish this.

With each breath detected, a five seconds amount is subtracted from the timer, instead of resetting it directly to zero. This is called Breath-Weighting. It avoids that a single - perhaps artificial - breath is resetting the timer to zero - and by doing so – delaying or omitting a “No RESP” alarm.

The counter is only reset to zero when two breaths are detected within five seconds. For example: With breathweighting, a single gasp 18 seconds into a No RESP (assuming a No RESP delay time of 20 seconds), will delay the detection of the alarm by five seconds; without breathweighting, the No RESP timer would have been reset to zero and the No RESP alarm would have been delayed by 18 seconds.

**Note:** Breath-Weighting may cause a “No RESP” alarm with a patient connected to a ventilator and/or a respiration rate less than 10 to 15 breaths per minute.

# ECG/Respiration Monitoring

Once the electrodes have been attached to the patient, proceed as follows:

- Connect the lead wires to the ECG patient cable, matching the colored end of the lead wire to the corresponding color on the cable.

**Warning:** Keep patient cables and lead wires away from the patient's neck area to avoid entanglement and accidental strangulation.

- Connect the round end of the ECG patient cable to the ECG/RESP connector on the side panel of the monitor.
- Press the POWER key to turn on the monitor.
- Select the appropriate ECG lead to be displayed on the Main screen. To optimize Respiration monitoring, Lead I configuration yields the most desirable results when the largest chest movements can be detected.
- If necessary, select the ECG and Respiration waveforms to be displayed on the Main Screen.
- Verify the signal quality with the help of the displayed ECG and Respiration waveforms.
- Wait for the monitor to determine the initial Heart Rate and Respiration Rate values. If the heart and lung visual indicators do not correspond to the patient's heart/respiration rate, reposition the electrodes until the indicators flash in synch with the patient's heartbeat and breathing. This will help to minimize false alarms.
- Check the Alarm Limits and configure them appropriately for the patient.

## Monitoring Pacemaker Patients

**Warning:** Keep pacemaker patients under close observation. Heart Rate determination may continue to count the pacemaker rate during cardiac arrest and some arrhythmias. Therefore, do not rely entirely on Hear Rate alarms.

**Warning:** False low heart rate indicators or false asystole calls may result with certain pacemakers because of electrical overshoots.

**Note:** ECG monitoring with patients on non-invasive transcutaneous pacemakers may not be possible due to large amounts of energy produced by these devices. Monitoring ECG with an external device may be needed.

To monitor a patient with a pacemaker, proceed as follows:

- Activate the pacemaker detection function as described in paragraph “Pacemaker Detection” on page 95.
- Follow institutional standards and change lead configuration as required.

When the monitor has detected a pacemaker pulse, it is rejected from being counted as a heartbeat. An artificial pulse is added to the ECG waveform display and the letter “P” is displayed to the right of the heartbeat icon to indicate the pacemaker discharge.

**Note:** Electrodes may need to be repositioned to modify detection of the electrical signals generated by a pacemaker.

## Disconnection of Lead Wires and Patient Cable

To remove lead wires from a patient cable, always grasp the strain relief plastic portion of the lead wire. Do not pull the wire itself.



Figure 15: Detaching the Lead Wires

To remove the patient cable from the monitor's ECG/RESP connector, press and hold down on the release button to unlock the cable and pull straight back. Do not pull on the cable itself.



Figure 16: Detaching the Patient Cable

## ECG/Respiration Display Window




Heart Rate (HR) and Respiration Rate (RR) are displayed in the green window at the top right area of the Main screen.

**Note:** When $CO_2$ is enabled, the RR numeric value will be smaller in size indication it being a secondary parameter.

## ECG/Respiration related Settings

Access the Parameters Menu as described on page 40 to review or change the following settings:

- ECG Lead selection
- ECG Trace Speed selection
- ECG Gain (sensitivity)
- Respiration Trace Speed selection
- Respiration selection (On or Off)

Access the Alarm Limits Menu as described on page 45 to review or change the following settings:

- ECG Alarm Limits
- Respiration Alarm Limits

**Note:** If either of the Heart Rate Alarm Limit values is set to OFF, an Alarm Off bell icon is displayed to the right of the heart beat icon.

Access the Audio/Visual Menu as described on page 41 to review or change the following settings:

- Beep Source (ECG or $SpO_2$)
- Beep Volume

Access the monitor configuration (refer to section “Setup” starting on page 91) to change the following settings:

- Pacer Detection selection (On or Off)
- CVA Filter selection (On or Off)

# ECG/Respiration Troubleshooting

## ECG/Respiration related Messages

The following ECG related messages may occur and be displayed in place of the ECG trace:

**`ECG Error`**

A communication failure occurred on the ECG Board. The monitor needs to be serviced.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.

**`ECG Module`**

A failure occurred on the ECG Board. The monitor needs to be serviced.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.

**`Leads OFF`**

The electrodes may be disconnected from the patient or dried out. A lead wire or the patient cable may be disconnected or broken.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.
- Reconnect electrodes to patient
- Use a new set of electrodes
- Make sure the lead wires are connected securely.
- Make sure the patient cable jack on the monitor is plugged in securely.
- Replace the lead wires or patient cable if necessary.

# 5 $SpO_2$

## Pulse Oximetry

## Introduction

The Pulse Oximeter parameter ($SpO_2$) determines arterial oxyhemoglobin saturation by measuring the absorption of red and infrared light passing through the tissue. Changes in absorption caused by pulsations of blood are used to determine Arterial Saturation (%$SpO_2$) and Pulse Rate (PR).

If $SpO_2$ was not selected as a trace to be displayed, a $SpO_2$ signal bar graph will be displayed indicating the relative signal strength and signal quality at the sensor site.

**Note:** The bar graph indicator is not proportional to the pulse volume.

An audio “beep” can be enabled that is generated each time the $SpO_2$ module detects a pulsation.

## Safety

**Warning:** If uncertain about the accuracy of any measurement, check the patient’s vital signs by alternate means and then make sure the monitor is functioning correctly.

**Warning:** Accurate oxygen saturation measurements cannot be obtained when the oximeter is not measuring the pulse properly. If the $SpO_2$ waveform, perfusion bar graph or the Pulse Rate be erratic or inaccurate, first examine the patient for any sign of distress and only then re-examine sensor placement.

**Warning:** A pulse oximeter should be considered an early warning device. As a trend toward patient deoxygenation is indicated, blood samples should be analyzed by a laboratory co-oximeter to completely understand the patient’s condition.

**Warning:** Various patient or mechanical conditions may cause inaccurate measurements. These conditions include but are not limited to: venous pulsations, hypotension, severe vasoconstriction, severe anemia, hypothermia, shock, cardiac arrest, sensor tension, sensor placement on the same extremity as a blood pressure cuff, arterial catheter or intravascular line, or arterial occlusion proximal to the sensor.

**Warning:** *MRI Scanning* - Do not use Nellcor oximetry sensors during magnetic resonance imaging (MRI) scanning. Conducted current could cause burns.

**Warning:** Do not expose probe detector to strong ambient light while monitoring a patient. A poor signal may result.

**Warning:** If a probe is damaged in anyway, discontinue use immediately.

**Caution:** Cardiogreen and other intravascular dyes, depending on the concentration, may affect the accuracy of the oximeter measurement.

**Caution:** Significant levels of dysfunctional hemoglobins such as carboxyhemoglobin or methemoglobin may affect the accuracy of the measurement.

**Caution:** Certain antibiotics, such as the Sulfas, can create high levels of methemoglobin. Methemoglobin is unable to bind $O_2$ and will absorb light similarly to reduced hemoglobin, thus giving an artificially low $SpO_2$.

**Caution:** *Reusable* $SpO_2$ sensors - Prolonged monitoring may require changing the probe site periodically. The sensor must be removed and repositioned every 4 hours and if indicated by circulatory condition or skin integrity, reapplied to a different monitoring site.

**Caution:** *Disposable $SpO_2$ sensors* - Prolonged monitoring may require changing the probe site periodically. The sensors must be removed and repositioned every 8 hours and if indicated by circulatory condition or skin integrity, reapplied to a different monitoring site. This sensor type should not be used with patients that exhibit allergic reactions to adhesive tape.

**Caution:** *Masimo LNOP probe* - Tissue damage can be caused by incorrect application or use of an LNOP probe, for example by wrapping the probe too tightly. Inspect the probe site as directed in the probe's directions for use to ensure skin integrity and correct positioning and adhesion of the probe.

## $SpO_2$ Sensors

The monitor can be equipped to use $SpO_2$ sensors manufactured by Masimo or Nellcor. The appropriate manufacturer's logo is shown next to the $SpO_2$ input connector.

**Note:** No other manufacturer's sensors should be used.

Select a sensor based on the patient size and monitoring conditions and properly attach the sensor to the patient.

**Note:** Consult instructions enclosed with each sensor for proper application.

There are a variety of $SpO_2$ sensors designs for a specific clinical application. Please refer to the Instructions of Use, which are supplied with each type of sensor.

## Preparations

Follow the instructions provided by the sensor manufacturer (Masimo or Nellcor) and prepare for $SpO_2$ monitoring as follows:

- Select a sensor that has the appropriate size for the patient's digit or extremity.
- Clean the surface of the probe before and after each patient use.
- Correctly position and attach the probe to the patient.
- Connect the sensor cable to the $SpO_2$ patient cable.
- Verify a secure connection and gently tug on the patient cable connector.
- Plug the $SpO_2$ patient cable into the $SpO_2$ connector on the left side panel of the monitor.

**Note:** The $SpO_2$ probe must be kept as motionless as possible to make a proper determination. To minimize motion artifacts, secure the sensor cable independently from the sensor. Tape may be used to secure the cable to the patient, e.g. around the base of the finger. Make sure that the tape being used does not restrict the blood flow.

## $SpO_2$ Monitoring

Once the sensor has been attached to the patient and the cable to the monitor, proceed as follows:

- Press the POWER key to turn the monitor on.
- If required, configure the appropriate waveform to be displayed on the Main Screen.
- Verify the signal quality and strength with the help of the $SpO_2$ waveform or the bar graph indicator.
- Wait for the monitor to determine the initial %$SpO_2$ and Pulse Rate values.

**Caution:** Some sensors may not be appropriate for a particular patient. If at least ten seconds of one bar pulses cannot be observed for a given sensor, change sensor location or sensor type until this condition is achieved.

- Check the alarm limits and configure them appropriately for the patient.

**Note:** If $SpO_2$ is selected as a waveform parameter, no bar graph indicator is displayed.

**Note:** Inspect the $SpO_2$ sensor site every 2 to 4 hours or per hospital protocol. If there is any skin irritation caused by the sensor, remove the sensor and apply it to a different location.

## SatSeconds™ Alarm Management

**Note:** The SatSeconds™ alarm management feature is only available in monitors with Nellcor's Oximax technology.

False or nuisance alarms are a common concern with pulse oximetry monitoring. Nuisance alarms are often triggered by minor and brief desaturation events that are clinically insignificant. Clinicians tend to manage these alarms by ignoring them, turning off the alarm or monitor, or widening the alarm limits. The CAS 750 Monitor incorporates Nellcor's SatSeconds Alarm Management feature that offers a better way to manage nuisance alarms without sacrificing patient safety.

The SatSeconds function can be activated from the Alarm Limits menu by selecting a SatSeconds limit, or "clock" of 10, 25, 50, 100 or OFF (0) SatSeconds. Clinicians who choose to employ the SatSeconds function should select a limit suited to their clinical environment and patient conditions. Think of SatSeconds as the product of magnitude and time a patient exceeds $SpO_2$ alarm limits. For example, 3 points below the alarm limit for 10 seconds equals 30 SatSeconds. An alarm is only triggered if a desaturation event occurs that reaches the SatSeconds limit you selected. As a safety net, when three or more $SpO_2$ alarm violations occur within 60 seconds, an alarm will sound even if the SatSeconds limit has not been reached.

To activate the SatSeconds function, proceed as follows:

- Press the LIMITS key to enter the Alarm Limits menu.
- Use the DOWN key to select "SatSeconds".
- Use the NEXT key to scroll to the next column to the right.
- Use the UP or DOWN keys to increase or decrease the SatSeconds value.
- Press the RETURN key when finished. All changes will be saved and become immediately effective.

**Note:** When SatSeconds is set to OFF, the alarm delay for %$SpO_2$ is based on the current configuration. Refer to paragraph "Selecting Alarm Delay" on page 95.

### Disconnecting $SpO_2$ Accessories

When $SpO_2$ monitoring is not required, disconnect the patient cable from the monitor by squeezing the tabs with your thumb and index finger while pulling the connector away from the monitor.

**Note:** To avoid damage to the Interface Cable, always hold it by the connector rather than the cable when connecting or disconnecting either end.

When the probe is disconnected from the monitor, an Equipment Alarm is activated and the message “No $SpO_2$ Probe” appears. Press the ALARMS key to acknowledge the alarm.

## $SpO_2$ Display Window




Oxygen Saturation (%$SpO_2$) and Pulse Rate (PR) are displayed in the yellow window at the lower right area of the Main screen.

**Note:** The bar graph indicator is only presented if the $SpO_2$ waveform is not displayed in one of the trace windows.

**Note:** When ECG is enabled, the PR numeric value will be smaller in size indication it being a secondary parameter.

## $SpO_2$ related Settings

Access the Parameters menu as described on page 40 to review or change the following settings:

- $SpO_2$ Trace selection
- $SpO_2$Trace Speed selection

Access the Alarm Limits menu as described on page 45 to review or change the following settings:

- $SpO_2$ Alarm Limits

Access the Audio/Visual menu as described on page 41 to review or change the following settings:

- Beep Source (ECG or $SpO_2$)
- Beep Volume

# $SpO_2$ Troubleshooting

**Note:** If the accuracy of any measurement does not seem reasonable, first check the patient's vital signs by an alternate method.

**Note:** The %$SpO_2$ and related Pulse Rate numeric displays will show dashes "- - -" and the signal strength window will be blank when no $SpO_2$ probe is attached to the monitor.

**Note:** Inaccurate $SpO_2$ measurements may be caused by:

- Anemia or low hemoglobin concentrations.
- Electro surgical interference.
- Excessive ambient light.
- Excessive patient movement.
- Incorrect sensor application or use.
- Intravascular dyes such as indocyanine green or methylene blue.
- Moisture in the sensor.
- Placement of a sensor on an extremity with a blood pressure cuff, arterial catheter, or intravascular line.
- Venous pulsations.

**Note:** The loss of a pulse signal can occur in any of the following situations:

- A blood pressure cuff is inflated on the same extremity as the one with the $SpO_2$ sensor attached.
- Excessive ambient light such as from a surgical lamp, a bilirubin lamp, or sunlight is present.
- The patient has hypotension, severe vasoconstriction, severe anemia, or hypothermia.
- The patient is in cardiac arrest or is in shock.
- The sensor is too tight.
- There is arterial occlusion proximal to the sensor.

## $SpO_2$ related Messages

**Note:** If $SpO_2$ is selected as a waveform parameter, messages appear in the waveform area of the display. If $SpO_2$ is selected as a numeric parameter, the messages appear in the Equipment Message Window of the display.

The following $SpO_2$ related messages may occur:

**`Chk SpO2 Probe`**

**Note:** This message relates to Masimo technology only.

The monitor is questioning the quality of the signal being received by the $SpO_2$ probe or the probe is receiving too much ambient light.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.
- Verify that the probe being used is the correct one for the monitor's $SpO_2$ configuration.
- Replace a wrong or defective probe.
- Verify that the probe is being used according to the manufacturer's recommendations.
- Verify that the probe emitter and detector are parallel to and directly opposing each other.

**`Low Perfusion`**

**Note:** This message relates to Masimo technology only.

The perfusion level being received by the $SpO_2$ probe is low.

- Determine if an extreme change in the patient's physiology and blood flow at the monitoring site occurred (e.g. an inflated blood pressure cuff, a squeezing motion).
- Try to warm the patient or the probe site.
- Move probe to a site with better perfusion.

**Low Signal IQ**

**Note:** This message relates to Masimo technology only.

The quality of the signal level being received by the $SpO_2$ probe is in question.

- Ensure proper probe type and application.
- Verify that the probe emitter and detector are parallel to and directly opposing each other.
- Clean or replace the probe.

**Note:** When the message "Low Perfusion" or "Low Signal IQ" appear and the monitor is displaying valid %$SpO_2$ numerics, no audible or visual alarms will be generated. In case the monitor is not displaying valid %$SpO_2$ numerics, an audible and visual alarm will be generated.

**No $SpO_2$ Probe**

The monitor is not detecting the $SpO_2$ probe. The probe was disconnected from either the Interface Cable or from the monitor.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.
- Reconnect probe to monitor.

**Probe Error**

The $SpO_2$ probe being used is not the correct one for the monitor's configuration.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.
- Verify the probe being used is the correct probe.

**Pulse Search**

The monitor is searching for a Pulse signal. This is normal at power-up as the monitor searches for a pulse or the probe position may have changed.

- Check the probe site.

**$SpO_2$ Error**

A failure occurred on the $SpO_2$ board. The monitor needs to be serviced.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.

**$SpO_2$ Module**

A failure occurred on the $SpO_2$ board. The monitor needs to be serviced.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.

**$SpO_2$ Probe OFF**

The monitor is no longer receiving a patient signal from the $SpO_2$ probe. The probe is no longer in contact with the patient.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.
- Check if probe is properly applied to the patient.
- Check if probe cable is connected to extension cable.
- Check if cable is attached to monitor.
- Replace probe and/or cable.

This page is intentionally left blank

# 6 NIBP

## Non-Invasive Blood Pressure

## Introduction

The Non-Invasive Blood Pressure (NIBP) parameter is using the oscillometric measurement technique. It automatically inflates an occluding cuff and then releases the cuff pressure step by step. It determines systolic, diastolic and mean arterial pressure. The NIBP function is suitable for neonatal, pediatric and adult patients.

## Safety

**Warning:** If uncertain about the accuracy of any measurement, check the patient's vital signs by alternate means and then make sure the monitor is functioning correctly.

**Warning:** Do not place the cuff on a limb being used for A-V fistulas, intravenous infusion or on any area where circulation is compromised or has the potential to be compromised.

**Warning:** When monitoring over an extended period of time, or at frequent intervals, periodically observe the patient's limb to make sure that the circulation is not impaired for a prolonged period of time.

**Warning:** The position of subject, physiological condition and other factors affect the NIBP readings. Blood pressure and pulse rate can fluctuate greatly between measurements; the monitor cannot alert the user to changes in vital signs occurring between measurement cycles.

**Warning:** Occasionally, electrical signals at the heart do not produce a peripheral pulse. If a patient's beat-to-beat pulse amplitude varies significantly (for example, pulsus alternans, atrial fibrillation, rapid-cycling artificial ventilator), blood pressure and pulse rate readings can be erratic and an alternate measuring method should be used for confirmation.

**Warning:** Do not apply external pressure against the cuff while monitoring. Doing so may cause inaccurate blood pressure values.

**Caution:** Always be sure to check that there are no signs of prolonged impairment of circulation regardless of automatic safety features that restrict the over all measurement time.

**Caution:** In shock conditions, the low amplitude of the blood pressure waveform may make it difficult for the monitor to accurately determine the systolic and diastolic pressures.

**Caution:** The monitor does not operate effectively if a patient is having seizure activity, tremors or is connected to a heart/lung machine.

**Caution:** Consult a physician for interpretation of blood pressure measurements.

**Caution:** The pulse rate derived from an NIBP measurement may differ from the heart rate derived from the ECG waveform because the NIBP parameter measures actual peripheral pulses, not electrical signals or contraction from the heart.

**Caution:** As with any non-invasive oscillometric blood pressure monitor, the accuracy of the measurements obtained may be adversely affected by the presence of agents, which alter the patient's cardiovascular system.

**Caution:** Do not operate the monitor's NIBP function unless it has been properly calibrated. Inaccurate blood pressure readings may result.

**Caution:** When a patient is experiencing arrhythmias during a measurement, the accuracy of the pulse determination may be affected or the time needed to complete a measurement may be extended.

**Caution:** Do not alter the monitor's air hose. CAS Medical Systems, Inc. cannot ensure proper monitor performance if the tubing is altered. Modification of the air hose will void the warranty. Avoid compression or restriction of pressure tubes.

**Note:** A patient's vital signs may vary dramatically during the use of cardiovascular agents such as those that raise or lower blood pressure or those that increase or decrease heart rate.

**Note:** The maximum amount of time allowed completing a NIBP measurement is 120 seconds in Adult mode and 90 seconds in Neonate mode. If the measurement has not been completed within that time, the cuff is deflated automatically and a message is displayed indicating the problem.

**Note:** To prevent exposure of the extremity to an inordinately high pressure, the cuff is deflated automatically when the pressure in the system is greater than 290 mmHg in the Adult mode or 145 mmHg in the Neonate mode.

**Note:** In the event of a microprocessor failure, the cuff will be deflated automatically within 10 seconds.

# Preparations

## Patient Mode Selection

**Caution:** Prior to patient monitoring, ensure the monitor is configured to the appropriate patient mode – Neonate or Adult. Refer to paragraph "Selecting Patient Mode" on page 92.

Select the Neonate mode when measuring on newborn patients and for all babies where the systolic pressure is expected to be below 120 mmHg. Use the Adult mode for infant and adult patients.

Switch from Neonate to Adult mode if one or more of the following conditions occur:

- Systolic pressure is expected to be above 120 mmHg.
- A cuff width of 6 cm or larger is required.

## NIBP Hose Selection

Two different inflation hoses are available.

- The coiled ten-foot NIBP hose is recommended for adult and pediatric patients.
- The straight six-foot NIBP hose is recommended for neonatal patients and small infants.

## Cuff Selection

The use of the correct size of cuff is essential for the accurate measurement of the blood pressure.

**Note:** CAS recommends the use of its reusable, disposable and neonatal cuffs.

**Width:** The widest cuff that can be placed around the upper arm or thigh should be used. A cuff that is too small for the arm will not supply sufficient pressure to the artery. This can cause an erroneously high blood pressure reading.

**Length:** The edge of the cuff is marked with a white arrow. It should fall within the white range marking when the cuff is applied to the patient. Otherwise the cuff is too long or too short.


TUFF CUFF
REF • Adult • CR5214
Bladder Size = 14 x 37.5 cm
CAS MEDICAL SYSTEMS, INC.
Branford, CT USA 06405
U.S. PATENT NO. 5,101,830
MAKE SURE THAT THIS ARROW
IS WITHIN THIS AREA
WHEN ON PATIENT.

Figure 17: Cuff Application Range Marker

**Note:** Overlapping the cuff will not affect the measurement results.

## Cuff Application

Measurements made above the level of the heart will give reduced blood pressure readings while measurements made below the heart level will give increased readings. These errors are mainly due to the weight of the blood.

Figure 18: Cuff Application

- When applying the cuff, make sure the cuff tubing is centered over the brachial artery.
- Wrap the cuff for a snug fit to avoid prolonged pumping time. Do not wrap a cuff over the patient's clothing.
- Verify the hook and loop sections of the cuff are fully engaged when it is wrapped around the limb.
- The limb should be positioned to be at heart level.
- Do not compress the cuff or the cuff hose. The hose must not be kinked or pinched.

**Warning:** The cuff should not be applied on a limb being used for an intravenous infusion. Do not place the cuff on any extremity being used for $SpO_2$ monitoring.

**Note:** Remember that there may be a marked difference between readings taken from the left and the right arms. Be consistent with each patient.

## NIBP Measurements

**Warning:** Before using the monitor on a new patient, select the appropriate patient's NIBP Initial Inflation cuff pressure.

**Warning:** When monitoring over an extended period of time, or at frequent intervals, periodically observe the patient's limb to make sure that the circulation is not impaired for a prolonged period of time.

**Caution:** Make sure the monitor is set to the correct Patient Mode. Refer to page 32 or to paragraph "Selecting Patient Mode" on page 92 to change the patient mode.

**Note:** The cuff target pressure must be higher than the patient's systolic pressure to obtain an accurate systolic and diastolic reading.

When the Initial Inflation Pressure (Init Inflate) is set to Adaptive, the initial cuff inflation pressure is 150 mmHg in Adult mode and 85 mmHg in Neonate mode. After the first measurement, the monitor will adapt the cuff inflation pressure automatically based on the previously measured systolic value (approximately 30 mmHg above). This will continue until the monitor is turned off and on again or the inflation pressure is selected manually.

When the Initial Inflation Pressure is set to any non-Adaptive setting, the initial cuff inflation pressure selection shall remain in affect until it is altered or when the Patient type is changed.

- To select an alternate Initial Inflation Pressure, enter the NIBP menu as described in this section on page 79.
- Select the desired "Init Inflate" pressure.

- Return to the Main Screen by pressing the RETURN key. The selected Initial Inflation Pressure will be displayed in the NIBP window.

**Note:** The Initial Inflation Pressure setting is saved during a power cycle.

**Caution:** When measuring blood pressure on a pediatric patient using the Adult mode, it is recommended that the Initial Inflation Pressure be set to a lower value (e.g. 120 mmHg).

Once the cuff is selected and applied to the patient, proceed with the following steps:

- Connect the cuff to the end of the NIBP hose and make sure the hose is connected to the NIBP connector on left side of the monitor.

## Starting a Blood Pressure Reading

**Caution:** Excessive motion can contribute to inaccurate measurements. It is important that the patient be kept still during a measurement. Make every attempt to alleviate patient's fear, anxiety and pain.

**Note:** The NIBP key will start a NIBP measurement regardless of the type of screen that is currently displayed.

- Press the NIBP key to start a measurement.

The cuff inflation pressure will be displayed until the measurement is completed.

The measurement typically takes less than 30 seconds to complete. In no case will the cuff remain pressurized for more than 120 seconds for Adult/Pediatric patients and no more than 90 seconds for Neonates.

When the measurement is completed, the cuff will automatically deflate, the monitor will provide an audio indication and display the NIBP values: Systolic/Diastolic and (MAP) or Systolic/Diastolic and Pulse Rate.

**Note:** If any displayed NIBP measurement were to be left on the screen for up to 24 hours, the monitor will automatically blank the NIBP displays to all dashes "- - -".

## Stopping a Blood Pressure Reading

A blood Pressure Reading can be interrupted for any reason at any time.

- Pressing the NIBP key during an ongoing NIBP measurement will interrupt the measurement and immediately deflate the cuff.

### Entering the Cycle Mode

If a cycle time is selected the monitor will automatically take blood pressure measurements at pre-selected time intervals.

- Enter the NIBP Menu as described in this section on page 79.
- Select the desired “NIBP Cycle” time.
- Select an alternate Initial Inflation Pressure if necessary.
- Return to the Main Screen by pressing the RETURN key. The selected cycle time will be displayed in the NIBP window.
- Press the NIBP key to begin the first measurement.

**Note:** If the Initial Inflation Pressure was altered, the first measurement will be made using that setting. After the first measurement, the monitor will switch to the Adaptive mode and adapt the cuff inflation pressure automatically based on the previously measured systolic value (approximately 30 mmHg above).

The measurement results are displayed in the NIBP Numeric Area of the Main Screen until the end of the next Automatic measurement or the start of the next Manual measurement. The remaining time until the next measurement will be displayed as well.

**Note:** If a measurement is desired between measurement cycles, press the NIBP key. After this measurement, the monitor will continue the Automatic Cycle mode.

**Note:** The Automatic Cycle mode is not saved during a power cycle.

### Terminating the Cycle Mode

To return to the Manual NIBP mode, proceed as follows:

- Enter the NIBP Menu as described in this section on page 79.
- Change the “NIBP Cycle” entry back to Manual.

### STAT Mode

The monitor will take a series of automatically started blood pressure measurements for a period of 5 minutes when STAT mode is activated. Each measurement is followed immediately by the next one after a 10 seconds pause that allows venous blood to return.

**Warning**: Readings obtained during STAT mode may not meet the stated accuracy of this monitor.

**Warning**: In some cases, rapid, prolonged cycling of an oscillometric, noninvasive blood pressure monitor cuff has been associated with any or all of the following: ischemia, purpura, or neuropathy. Apply the oscillometric cuff appropriately, according to instructions, and check the cuff site and cuffed extremity regularly when blood pressure is measured at frequent intervals or over extended periods of time.

### Entering STAT Mode

To activate the STAT mode, proceed as follows:

- Enter the NIBP Menu as described in this section on page 79.

**Note:** The monitor must be in the Manual NIBP Cycle mode when selecting STAT mode.

- Select an alternate Initial Inflation Pressure if necessary.
- Move the cursor to STAT and then select ON.
- Return to the Main Screen by pressing the RETURN key. The message “STAT” appears in the NIBP window.
- Press the NIBP key to begin the first measurement.

**Note:** If the Initial Inflation Pressure was altered, the first measurement will be made using that setting. After the first measurement, the monitor will switch to the Adaptive mode and adapt the cuff inflation pressure automatically based on the previously measured systolic value (approximately 30 mmHg above).

Between readings, the NIBP window will display the remaining seconds until the next measurement.

**Note:** Selecting STAT mode will override any time interval selected for the NIBP measurement cycle.

### Exiting STAT Mode

The monitor will stop the STAT mode automatically after 5 minutes and returns to the Manual NIBP mode.

- Press the NIBP key to terminate the current NIBP measurement.
- Enter the NIBP Menu as described in this section on page 79.
- Select STAT mode OFF.

# NIBP Display Window

The NIBP numeric values (SYS, DIA, MAP and/or Pulse Rate) are displayed in the blue window at the lower center of the Main screen.

# NIBP Menu

### To enter the NIBP Menu

Press and hold the NIBP key for 2 seconds to access the NIBP menu. The following screen will be displayed:

| NIBP | |
|---|---|
| NIBP Cycle | Manual |
| Init Inflate | Adaptive |
| STAT | OFF |
| NIBP Numerics | S/D (MAP) |

Figure 19: NIBP Menu

**Note:** While in the NIBP menu and no key is pressed for 30 seconds, the monitor will automatically save all changes made and return to the Main Screen.

### Menu Options

The NIBP menu allows the user to:

- Select the NIBP Cycle time. Available selections are: Manual, 1, 2, 3, 4, 5, 10, 15, 30, 60 or 90 Minutes.
- Select the Initial Inflation Pressure for the next NIBP measurement. Available selections are:

  Neonate mode: Adaptive, 60, 80, 85, 100 or 120 mmHg
  Adult mode: Adaptive, 80, 100, 120, 140, 150, 160, 180 or 200 mmHg
- Select the STAT mode to be OFF or ON
- Select how the NIBP numerics are to be displayed in the NIBP display window. Available selections are: S/D (MAP), S/D (PR) or (MAP) S/D.

### Operating the NIBP Menu

To change settings, proceed as follows:

- Use the UP and DOWN keys to select the desired row.
- Move the cursor to the right with the use of the NEXT keys.
- Use the UP and DOWN keys to select the desired function or setting.

## NIBP related Settings

Access the Parameters menu as described on page 40 to review or change the following settings:

- NIBP History selection for Trace 3.

Access the Alarm Limits menu as described on page 45 to review or change the following settings:

- NIBP Alarm limits

# NIBP Troubleshooting

**Note:** If the accuracy of any measurement does not seem reasonable, first check the patient's vital signs by an alternate method.

## NIBP related Messages

- Pressing the ALARMS key will cause the following NIBP related messages to be removed from the display and silence the audio alarm tone.

**Air Leak**

Air leak in the cuff, the NIBP hose or in the internal pneumatic system of the monitor.

- Check that the cuff/hose/monitor connection is secure.
- Check cuff for leaks. Do not use a known leaky cuff.

**Appl Error**

Neonate cuff is detected in Adult Mode.

- Check cuff. Replace cuff or change operating mode.

**Flow Error**

Stable cuff pressure cannot be maintained by the NIBP pneumatic system.

- Check the external tube for kinks.
- Perform a Pneumatic Pressure Check as described on page 108.
- Replace cuff.

**Loose Cuff**

Cuff applied too loosely.

- Check cuff for proper fit on patient.

**Motion**

There was too much extremity motion for the monitor to accurately complete the NIBP measurement in 120 seconds. Measurements can be obtained when there is limited extremity movement, but the measurement time may be extended.

- Restrain patient extremity motion.

**NIBP Cal**

Pressure calibration data corrupted within NIBP module. Pressure module needs recalibration. The monitor needs to be serviced.

**NIBP Error**

A failure occurred on the NIBP board. The monitor needs to be serviced.

**NIBP Module**

A failure occurred on the NIBP board. The monitor needs to be serviced.

**Over Press**

Cuff pressure exceeded 290 mmHg in the Adult mode or 145 mmHg in the Neonate mode. Very rapid squeezing of the cuff can cause this error.

- Repeat the measurement.
- If this message repeatedly occurs during normal use, the monitor must be serviced.

**Range Error**

The systolic reading exceeds the measurement range of 255 mmHg in the Adult mode or 135 mmHg in the Neonate mode.

- Repeat measurement.
- If the message is displayed again, use another method to measure the patient's blood pressure.

**Signal Sat**

Signal Saturation or motion pulses too strong.

- Repeat measurement. Limit patient activity; the arm must be still and/or relaxed.

**Time Out**

The monitor was unable to complete a measurement within 120 seconds in the Adult mode or 90 seconds in the Neonate mode. An extremely long measurement can be due to a loose cuff, high blood pressure, or monitor re-pumps.

- Repeat measurement.
- Try higher initial cuff pressure.
- If message consistently reappears try using another means to obtain patient's blood pressure.

**Weak Signal**

The monitor did not detect any pulses during a NIBP measurement.

- Check the fit of the cuff.
- Repeat measurement.

# 7 $CO_2$
## Capnography

## Introduction

The monitor uses Microstream non–dispersive infrared (NDIR) spectroscopy to continuously measure the amount of $CO_2$ during every breath. The amount of $CO_2$ at the end of exhalation ($EtCO_2$) and the Respiratory Rate (RR) are measured.

The $CO_2$ measurement is intended for use in any environment where continuous, noninvasive monitoring is desired, including hospital and mobile use (when protected from excessive moisture such as direct rainfall). It is intended for use on adult, pediatric, and infant/neonatal patients. It is intended for use on intubated and non-intubated patients.

A $CO_2$ library can be selected for educational viewing. Refer to paragraph "$CO_2$ Library" on page 41.

## Safety

**Warning:** If uncertain about the accuracy of any measurement, check the patient's vital signs by alternate means and then make sure the monitor is functioning correctly.

**Warning:** When monitoring $CO_2$ in the presence of flammable gases or anesthetics, such as high concentrations of oxygen or nitrous oxide, connect the gas outlet to a scavenger system.

**Warning:** $CO_2$ readings and respiratory rate can be affected by certain ambient environmental conditions, sensor application errors, and certain patient conditions.

**Caution:** Microstream consumables are latex free, non-sterile and non-pyrogenic.

**Caution:** During MRI scanning, the monitor must be placed outside the MRI suite. When the monitor is used outside the MRI suite, $CO_2$ monitoring can be implemented using the FilterLine XL.

# Preparations

**Caution:** Prior to patient monitoring, ensure the monitor is configured to the appropriate patient mode – Neonate or Adult. Refer to paragraph "Selecting Patient Mode" on page 92.

## Microstream $CO_2$ Consumables

**Note:** The Microstream $CO_2$ technology requires the use of Microstream consumables. No other manufacturer's products should be used.

When choosing $CO_2$ consumables for a particular patient, the following should be considered:

- Intubated or non–intubated.
- Whether the patient is on mechanical ventilation.
- Duration of use.
- Patient's size and weight

**Caution:** Before use, carefully read the Directions for Use that is enclosed with the Microstream consumable.

**Caution:** Microstream consumables are designed for single patient use, and are not to be reprocessed.

**Warning:** Do not attempt to disinfect or flush the FilterLine as the monitor can be damaged.

## Non-Intubated Application

For non-intubated patients the continuous gas sampling is taken from the patient's nose and/or mouth. The following consumables are available:

- CapnoLine: Cannula with nasal patient adapter.
- Smart CapnoLine Oral Nasal Cannula – for use in procedural sedation. Also available with $O_2$ delivery.
- CapnoLine H – for use in ICU for patients receiving hi–flow oxygen by mask or on long term CPAP or Bi-PAP. Also available with $O_2$ delivery.
- NIV–Line – for use under oxygen CPAP, Bi–PAP or NPPV mask and to avoid intubation during EMS transport.

Select the appropriate CapnoLine and connect it to the monitor before attaching it to the patient's airway. Be sure to follow Microstream $CO_2$ Consumables' *Directions for Use* for proper connection.

**Caution:** Loose or damaged connections may compromise ventilation or cause an inaccurate measurement of respiratory gases. Securely connect all components and check connections for leaks.

**Warning:** Carefully route the CapnoLine to reduce the possibility of patient entanglement or strangulation.

**Note:** Ensure the CapnoLine is not twisted or crimped and the cannula tips are in the nostrils.

**Note:** If the CapnoLine includes an additional oxygen connection, connect the oxygen connector to the standard outlet.

### Intubated Applications

For intubated patients the continuous gas sampling is taken from the Airway Adapter directly at the intubation tube. The following consumables are available:

- FilterLine set for non-humid environments.
- FilterLine H set for humid environments.

Select the appropriate FilterLine and connect it to the monitor before attaching it to the patient's airway. Be sure to follow Microstream $CO_2$ Consumables' *Directions for Use* for proper connection.

**Warning:** When used with closed suction system, do not place the airway adapter between the suction catheter and endotracheal tube.

**Caution:** Loose or damaged connections may compromise ventilation or cause an inaccurate measurement of respiratory gases. Securely connect all components and check connections for leaks.

**Note:** Ensure the consumable is not twisted or crimped.

**Note:** During nebulization or suction (when not using closed suction system), in order to avoid moisture buildup and FilterLine occlusion, disconnect the Airway Adapter from the patient's endotracheal tube.

## $CO_2$ Monitoring

Once the appropriate consumable has been applied to the patient, proceed as follows:

- Press the POWER key to turn the monitor on.
- If necessary, select the $CO_2$ waveform to be displayed on the Main screen.

**Note:** If $CO_2$ was not selected as a trace to be displayed, a $CO_2$ signal bar graph will be displayed indicating the relative signal strength and signal quality.

**Note:** While the $CO_2$ module is initializing (typically 30 to 40 seconds) the monitor shall display the message "$CO_2$ Warm-Up" and dashes "- - -" will remain in the numeric section.

- Check that the monitor is accurately detecting the $CO_2$ signal and wait for the monitor to determine the initial $EtCO_2$ and Respiration Rate values.
- If the lung visual indicator is not corresponding to the patient's respiration, reposition the sensor until the indicators flash in synch with the patient's breathing. This will help to minimize false alarms.

Check the alarm limits and configure them appropriately for the patient. Refer to paragraph "Alarm Limits Menu" on page 45.

### Removing the $CO_2$ Consumables

When $CO_2$ monitoring is no longer required, disconnect the consumable by carefully removing the connector from the $CO_2$ input receptacle.

When the consumable is disconnected from the monitor, the message "$EtCO_2$ OFF" is displayed, the Equipment Alarm LEDs are flashing and an audible alarm sounds, indicating a connection has been broken.

**Note:** If $CO_2$ is selected as a waveform, messages appear in the $CO_2$ waveform window - otherwise messages appear in the Equipment Message Window of the Main screen.

- To acknowledge the alarm, press the ALARMS key.

**Caution:** Dispose of Microstream consumables according to standard operating procedures or local regulations for the disposal of contaminated medical waste.

## $CO_2$ Display Window

End Tidal $CO_2$ ($EtCO_2$) and Respiration Rate (PR) are displayed in the yellow window at the lower left area of the Main screen.

**Note:** The bar graph indicator is only presented if the $CO_2$ waveform is not displayed in one of the trace windows.

# $CO_2$ related Settings

Access the Parameters menu as described on page 40 to review or change the following settings:

- $EtCO_2$ Trace selection
- $EtCO_2$ Trace speed selection
- $EtCO_2$ Scale selection
- $EtCO_2$ Print selection
- $EtCO_2$ Trace Type selection

Access the Alarm Limits menu as described on page 45 to review or change the following settings:

- $EtCO_2$ Alarm Limits

Access the Monitor Setup menu as described on page 96 to change the following settings:

- $EtCO_2$ Unit selection (mmHg, kPa or %).

# $CO_2$ Troubleshooting

**Note:** If the accuracy of any measurement does not seem reasonable, first check the patient's vital signs by an alternate method.

## $CO_2$ related messages

- Pressing the ALARMS key will silence the audio alarm tone for the following $CO_2$ related messages.

**`Blocked Line`**

The $CO_2$ module has attempted to purge the line with no success.

- Disconnect the FilterLine.
- Check the FilterLine for kinks and/or blockages.

**`Cal EtCO2`**

The $CO_2$ module's calibration timer has reached zero hours.

- Perform a $CO_2$ calibration check.

**`Clearing Line`**

The $CO_2$ module has detected an occlusion and is attempting to purge the line.

- Check the FilterLine tube for kinks and/or blockages.

**Chk $EtCO_2$ Flow**

The $CO_2$ module has detected a blockage in the exhaust port.

- Check the exhaust port and associated tubing for kinks and blockages.

**$EtCO_2$ Error**

A failure occurred on the $CO_2$ Board.

- To reset the $CO_2$ board turn the monitor off and disconnect it from external power (AC Line or 12VDC). Remove the battery, and then reconnect the battery and external power. Turn the monitor on.
- If the message remains, the monitor needs to be serviced.

**$EtCO_2$ Module**

A failure occurred on the $CO_2$ Board. The monitor needs to be serviced.

**$EtCO_2$ OFF**

The $CO_2$ module is no longer receiving a signal from the patient or the FilterLine is no longer connected to the monitor. The consumable is no longer connected to the monitor.

- Press the ALARMS key. The monitor silences the audio alarm tone, but the visual messages remain.
- Check the external tubing for disconnections.

**Service $EtCO_2$**

The $CO_2$ module's service timer has reached zero hours. The monitor will continue to function.

- Arrange for service as soon as possible.

# 8 TEMP

## Temperature

## Introduction

The monitor is designed to work with Yellow Springs (YSI) Series 400 temperature probes.

- Core Temperature probes.
- Skin Temperature probes.
- Esphogeal/Rectal Temperature probes.

## Safety

**Warning:** If uncertain about the accuracy of any measurement, check the patient's vital signs by alternate means and then make sure the monitor is functioning correctly.

## Preparations

**Note:** Consult instructions enclosed with each sensor for proper application.

Refer to section "Temperature Accessories" on page 149 for Temperature probe types and order information.

## Temperature Monitoring

To take a temperature measurement:

- Connect the probe to the Temperature Adapter cable.
- Connect the Adapter cable to the TEMP connector on the left side of the monitor.
- Press the POWER key to turn the monitor on.
- Check that the monitor is accurately measuring temperature when the value has been determined.

## Temperature Display Window

The Temperature value (TEMP) is displayed in the yellow window at the lower right area of the Main screen.

**Note:** The temperature unit may be °F or °C depending on its configuration.

**Note:** When the Temperature value being measured is greater than 43 °C (109.4 °F), the TEMP numeric value will be displayed as ">>>>>" indicating a value above the monitors temperature range.

**Note:** When no Temperature probe is connected, the numeric display for Temperature will be all dashes "- - -".

## Temperature related Settings

Refer to paragraph "Selecting Temperature Units" on page 96 to change the following settings:

- Temperature Unit selection (°F or °C).

## Temperature Troubleshooting

### Temperature related Messages

None

# 9 Setup

## Configuring the Monitor

## Entering the Setup Menu

The Setup menu shows all configuration options in a tabular form. To enter the Setup menu, proceed as follows:

- Press and hold the LIMITS and AUDIO/VISUAL keys while the monitor is being turned on.




Figure 20: Setup Menu

The cursor is automatically positioned at the first item in the left column.

- Use the UP and DOWN keys to select the item that needs to be changed.
- Use the NEXT key to move the cursor to the right column.
- Use the UP and DOWN key to scroll through the available options.
- Press the RETURN key to exit the Setup menu.

**Note:** If no key is pressed for 60 seconds, the monitor automatically exits the Setup menu.

**Note:** Depending on the monitor configuration, a scroll bar will be displayed indicating that more selections are available to view.

All changes will be saved and the monitor returns to the Main screen.

## Selecting the Language

To select the monitor's language, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select the desired language.

**Note:** Once a new language is selected, all text in the Setup menu (and in all other menus and screens) will be displayed in that language. This may cause difficulties if an unknown language is selected.

## Selecting Patient Mode

**Caution:** Prior to patient monitoring, ensure the monitor is configured to the appropriate Patient mode.

To select the Patient mode, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "Patient" is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select Adult or Neonate mode.

Or alternatively:

- Keep the UP key pressed when the monitor is turned on to set the monitor to Adult mode.
- Keep the DOWN key pressed when the monitor is turned on to set the monitor to Neonate mode.

The Patient mode will be displayed in the Equipment Message Window in the lower left area of the Main screen.

**Note:** ECG, $CO_2$, NIBP and Respiration functions are affected by selecting Adult or Neonatal operating modes. The factory default is the Adult mode.

## Configure Audio Silencing

The audio signal associated with patient alarms can be silenced by pressing the ALARM key. The functionality of this key depends on the selected configuration:

To configure the monitor's silencing function, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "Audio Silence" is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select the desired option.

Select "2 Minute" when the silence period should automatically terminate after 2 minutes.

Select "Permanent" when the silence period should continue permanently until it is terminated manually.

**Warning:** In Permanent Silence mode the user is responsible to reactivating (enabling) the audio alarm signal by pressing the ALARMS key again.

## Setting the Date

To set the monitor's Date function, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "Date" is selected.
- Press one of the NEXT keys to move the cursor to Day-Month-Year.
- Use the UP or DOWN keys to increase or decrease the values for Day-Month-Year.

**Note:** Altering the monitors' date will affect the History data.

## Setting the Time

To set the monitor's Time function, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "Time" is selected.
- Press one of the NEXT keys to move the cursor to Hours **:** Minutes.
- Use the UP or DOWN keys to increase or decrease the values for hours or minutes.

**Note:** Altering the monitors' time will affect the History data.

## Daylight Saving Time Option

To set the monitor's Daylight Saving Time function, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "DST" is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select on of the Daylight Saving Time options.

The possible options have the following meaning:

**OFF:** Daylight Saving Time is not activated. The monitor will not change the time automatically. It is up to the user for changing the time if needed. This is the monitor's default setting.

**N AMERICA**: Use this setting and the monitor will automatically change time accordingly to the Daylight Saving Time for countries in North America.

**EU 01:00**: Daylight Saving Time for all European countries using Greenwich Mean Time.

**EU 02:00:** Daylight Saving Time for all European countries 1 hour ahead of Greenwich Mean Time.

**EU 03:00:** Daylight Saving Time for all European countries 2 hours ahead of Greenwich Mean Time.

**Note:** Enabling Daylight Saving Time will affect the History data.

## Pacemaker Detection

To change the Pacemaker setting, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter “Pacer Detect” is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select OFF or ON.

## Selecting CVA Filter

To change the CVA Filter setting, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter “CVA Filter” is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select OFF or ON.

## Selecting Alarm Delay

To change the $SpO_2$ Alarm Delay setting, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter “$O_2$ Alarm Delay” is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select 10 Sec or 0 Sec for the delay.

## Selecting $CO_2$ Units

To select the display unit for $CO_2$ waveform and $EtCO_2$ values, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "$EtCO_2$ Unit" is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select one of the options (mmHg, kPa or %).

## Selecting Temperature Units

To select the display unit for Temperature values, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "TEMP Unit" is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select ºC or ºF.

## Selecting Monitor Set-Up Default

To select a preset set-up configuration mode, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter "Restore Default" is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select one of the options (None, Factory or Sleep Study).

**Note:** When configured for Sleep Study, the "Auto Dim" shall be set to On and the "Audio Silence" shall be set to Permanent.

Refer to the paragraphs "Auto Dim" on page 42 and "Configure Audio Silencing" on page 93 for more information.

## Selecting Display Background

To select the display background, proceed as follows:

- Enter the Setup menu as described on page 91 and follow the general instructions given there.
- Press the DOWN key until the parameter “Background” is selected.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to select Dark or Light.

This page is intentionally left blank

# 10 Cleaning

**Warning:** Do not, under any circumstances, perform any cleaning while the monitor is being used to monitor a patient. The monitor must be turned off. Unplug the monitor from the AC or DC power source and disconnect all accessories.

**Caution:** Do not open the monitor to clean it.

**Caution:** Do not immerse any part of the electrical connector of the cable or accessories in the cleaning or disinfection solution at any time. This may cause internal damage and reduce the product life.

**Caution:** Do not use abrasive cleaners, isopropyl alcohol or organic solvent for cleaning. Use of these cleaners can cause damage, stiffness and brittleness to the monitors' surface and to cables and wires.

## Cleaning the Monitor

**Caution:** Disconnect all accessories from the monitor before cleaning.

Examine the monitor's case for damages and check the AC power cord for bent or broken prongs, cracks or fraying. Neither the monitor nor the power cord should be used if damaged. If any damage is noted, contact the appropriate service personnel.

**Caution:** Do not spray or pour any water or cleaning solution directly onto the monitor.

- **Housing:** As needed, clean the monitor using a soft cloth dampened with a mild dishwashing detergent solution and gently rub the soiled area until clean. Use a clean soft cloth to dry the monitor.
- **Display:** Clean the display window using a soft, lint-free cloth sprayed with an alcohol free glass cleaner. The use of paper towels is not recommended as it may scratch the surface.
- **Disinfections:** The monitor surfaces may be disinfected using a soft cloth saturated with a 10% (1:10) solution of chlorine bleach in tap water. When all of the surfaces have been disinfected, wipe the entire surface of the monitor using a soft cloth dampened with fresh water to remove any trace amounts of residue and/or fumes.

**Note:** Thoroughly wipe off any excess cleaning solutions. Care should be taken to prevent water or cleaning solution to run into connector openings or crevices.

## Cleaning Patient Cable and Leadwires

Prior to each patient use, inspect the patient cable and lead wires for damage.

- Clean the patient cable and lead wires using a soft cloth dampened with a germicidal solution.

## Cleaning Cuffs and Pneumatic Hoses

**Note:** Refer to the product packaging for additional Cleaning and Disinfecting Instructions were applicable.

Prior to each patient use, inspect the blood pressure cuff and the pneumatic hose for damage.

- **Tuff-Cuff:** Clean the cuff using a soft cloth dampened with a 70% Isopropyl Alcohol solution. The cuff should be allowed to thoroughly dry before use.
- **Safe-Cuff:** Clean the cuff using a soft cloth dampened with a soap, water-based detergent or chlorinated disinfectant solution. Do not use alcohol.
- **SoftCheck:** These cuffs are designed for single patient use, and are not to be reprocessed.
- **UltraCheck:** For normal cleaning with mild detergents / dilute bleach solution (1-2%), wipe the cuff with the cleaning solution, rinse with water and dry.
- **Pneumatic Hose:** Clean the pneumatic tubing using a soft cloth dampened with a germicidal solution.

**Caution:** If the cuff or hose should become grossly contaminated with blood or other bodily fluids, it should be discarded.

**Note:** It is not recommend submersing the cuff or hose. Liquid should not be permitted to enter the cuff bladder or hose because instrument damage may occur. Cuffs and hoses should be allowed to thoroughly dry before use.

## Cleaning the $SpO_2$ Interconnect Cable

Prior to each patient use, inspect the $SpO_2$ Interconnect cable for damage.

- Clean the cable using a soft cloth dampened with a germicidal solution.

## Cleaning $SpO_2$ Sensors

**Caution:** Do not soak or immerse the sensor or its cable in any liquid solution. Do not attempt to sterilize it.

Refer to the Directions For Use pamphlet enclosed with each sensor for more information.

- Clean the $SpO_2$ sensors with a soft cloth dampened with 70% Isopropyl alcohol solution. Allow the sensor to dry prior to placement on a patient.

## Cleaning Temperature Probes

Refer to the Directions For Use pamphlet enclosed with each temperature probe for more information.

- Cleaned probe with a soft cloth dampened in mild detergent.

Disinfect probe if necessary:

- Use a soft cloth saturated with a 10% solution of chlorine bleach in tap water or 70% isopropyl alcohol.
- Wipe the entire surface with a soft cloth dampened with fresh water to remove any trace amounts of residue and/or fumes.

## Cleaning $CO_2$ Consumables

Microstream $CO_2$ consumables are designed for single patient use and are not to be reprocessed.

**Caution:** Do not attempt to disinfect or flush the FilterLine as the monitor can be damaged.

## Cleaning the Printer

Refer to the printer User's Manual for more information.

**Caution:** Before cleaning the printer, disconnect the AC adapter.

- Wipe with a soft dry cloth.

To remove extreme dirt buildup:

- Soak a cloth with mild detergent, wring well and wipe. Dry by wiping with a soft dry cloth.

**Caution:** Do not use volatile chemicals such as thinner, benzine, etc.

**Caution:** Never wet the inside of the printer mechanism.

This page is intentionally left blank

# 11 Maintenance

**Warning:** Do not, under any circumstances, perform any testing or maintenance on the monitor while the monitor is being used to monitor a patient. The monitor must be turned off. Unplug the monitor from the AC or DC power source and remove the internal battery.

**Caution:** Service performed by unauthorized personnel could be damaging to the monitor and may void the warranty. For service, contact your CAS Representative.

## Maintenance Intervals

Preventive maintenance of the monitor is an important function that should be performed routinely to ensure safe and efficient monitor operation. The following maintenance intervals are recommended:

- **Battery Pack:** Replace the battery every 2 years.
- **ECG/Respiration:** No user calibration is required. Perform a bench Test once a year or when there is doubt about the validity of the ECG and/or Respiration readings.
- **$SpO_2$:** No user calibration is required. Perform a $SpO_2$ Simulator Check once per year or when there is doubt about the validity of the $SpO_2$ readings.
- **NIBP:** Perform a Pneumatic Pressure check every 6 months and a NIBP Calibration Check once a year or when there is doubt about the validity of the pressure readings.
- **Temperature:** No user calibration is required. Perform a Temperature Calibration Check once per year or when there is a doubt about the validity of the temperature readings.
- **$CO_2$:** Perform a $CO_2$ Calibration Check once per year, after 4000 operating hours (whichever comes first) or when there is doubt about the validity of the $CO_2$ readings. The monitor should be returned to CAS Medical Systems for periodic maintenance every 14,000 operating hours.

# Maintenance Checks

## Entering the Service Menu

**Warning:** Do not enter the SERVICE menu when the monitor is connected to a patient.

To enter the monitor's SERVICE menu, proceed as follows:

- Press and hold the PARAMETERS and HISTORY keys while the monitor is being turned on.

| Service | |
|---|---|
| IrDA Test | OFF |
| $EtCO_2$ Check | OFF |
| $EtCO_2$ Cal | OFF |
| $EtCO_2$ CAL Date | 06-Feb-04 |
| $EtCO_2$ S/N | 00234 |
| Service $EtCO_2$ | xxxxx Hours |
| Cal $EtCO_2$ | xxxx Hours |
| Manometer Mode | OFF |
| Pneumatic Test | OFF |
| PIC Voltage | xx.xx v |

Figure 21: Service Menu

The cursor is initially positioned at the first item in the left column.

- Use the UP or DOWN keys to select any other item.
- Press one of the NEXT keys to move the cursor to the next column to the right.
- Use the UP or DOWN keys to scroll through the available options.

## Exiting the Service Menu

- Press the RETURN key to exit the menu and return to the Main screen.

**Note:** If no key is pressed for 15 minutes, the monitor will automatically return to the Main screen.

## IrDA Test

The IrDA Test checks the alignment of the monitor and the printer to guarantee communication between their infrared windows.

To use the Infrared Test, proceed as follows:

- Enter the Service menu as described on page 104 and follow the general instructions given there.
- Change the IrDA Test setting from “OFF” to “ON”.

The message “Printing” will be displayed in the Equipment Message Window. When the Infrared windows are properly positioned, the printer will begin to print a series of “A” characters. The test will continue for 60 seconds or can be terminated by the user.

- Press one of the PREVIOUS keys to cancel the test and return to the IrDA TEST menu selection.

## $CO_2$ Calibration Check

The $CO_2$ Calibration Check is a tool to verify proper calibration of the $CO_2$ parameter.

**Caution:** Do not perform a $CO_2$ Calibration Check when the monitor is in normal measuring mode. This mode corrects the $CO_2$ value for BTPS (Body, Temperature, Pressure, Saturation), which assumes that alveolar gases are saturated with water vapor. The Calibration Check mode disables this correction.

**Caution:** The Calibration Check must be performed with a manufacturer authorized Calibration Kit containing (5% $CO_2$, 21% $O_2$, Balance $N_2$) gas, tubing adapter and a calibration filter line. A manufacturer approved Calibration Kit can be purchased from Scott Medical (P/N 0304653ORFBD).

**Note:** The Calibration Check should be performed only after the $CO_2$ module has been operating for at least 20 minutes in a normal operating mode.

To check $CO_2$ Calibration, proceed as follows:

- Verify that the FilterLine supplied with the Calibration kit is firmly attached to the gas canister.
- Enter the Service menu as described on page 104 and follow the general instructions given there.
- Press the DOWN key until the parameter "$EtCO_2$ Check" is selected.
- Press one of the NEXT keys to select the "OFF" setting.
- Change the $EtCO_2$ Check setting from "OFF" to "Connect Gas".
- Connect the FilterLine to the monitor.
- The highlighted section should read "Open 5% Gas".
- When the message "Open 5% Gas" appears, press and hold open for 15 seconds the gas valve until the reading stabilizes.
- Verify the reading on the display to be 5.0% ± 0.3 vol%.

**Note:** Calibration is not required if the measured value are the same as the concentration of the calibration gas ± 0.3 vol%.

- Press one of the PREVIOUS keys to cancel the test and return to the $EtCO_2$ Check menu selection.
- Disconnect the FilterLine from the monitor.

Should the monitor fail the $EtCO_2$ Check, it is recommended to perform an $EtCO_2$ Calibration or returning the monitor for service.

## $CO_2$ Calibration

**Note:** Perform the Calibration only after performing the Calibration Check.

**Caution:** The Calibration must be performed with a manufacturer authorized Calibration Kit containing (5% $CO_2$, 21% $O_2$, Balance $N_2$) gas, tubing adapter and a calibration FilterLine. A manufacturer approved Calibration Kit can be purchased from Scott Medical (P/N 0304653ORFBD).

To perform a $CO_2$ Calibration, proceed as follows:

- Verify that the FilterLine supplied with the Calibration Kit is firmly attached to the gas canister.
- Enter the Service menu as described on page 104 and follow the general instructions given there.
- Press the DOWN key until the parameter "Et$CO_2$ Cal" is selected.
- Press one of the NEXT keys to select the "OFF" setting.
- Change the Et$CO_2$ Cal setting from "OFF" to "Connect Gas".
- Connect the FilterLine to the monitor.
- The highlighted section should read "Open 5% Gas".
- When the message "Open 5% Gas" appears, press and hold the gas valve until the message "Remove Gas" is displayed.
- The highlighted section should read "Passed" and two audible beeps should be heard when the calibration is completed.
- Verify the "Et$CO_2$ CAL Date" and the "Cal Et$CO_2$ Hours" have been updated.

The starting hours for Calibration are initially 1200 hours, and then 4000 after that. This timer is automatically reset to 4000 hours during the Calibration process as long as it has been more than 720 hours from the last calibration. If this was less than 720 hours, the timer does not reset.

- Press one of the PREVIOUS keys to cancel the test and return to the Et$CO_2$ Cal menu selection.
- Disconnect the FilterLine from the monitor.

**Note:** Should the monitor fail the Et$CO_2$ Calibration, it is recommended the procedure be repeated. Prior to repeating the Calibration procedure, carefully check all connections.

Should the monitor continue not to pass the Calibration procedure, the monitor should be returned for service.

## NIBP Checks

To guarantee correct NIBP measurements, it is recommended to perform the following checks at least once per year. All checks can be done in one session beginning with the Pneumatic Pressure Check, then continue with the Calibration Check and finally perform the Over Pressure Test.

**Note:** The monitor must be in the Adult mode to perform the following pressure checks.

The following equipment is required to perform the NIBP pressure checks:

- A Calibration Kit (P9) which is included with the monitor.
- A Mercury manometer or equivalent whose accuracy meets the AAMI/ANSI Standard for Non-Automated Sphygmomanometers, 2002.

**Warning:** Do not activate the Service menu when a cuff is attached to a patient.

### Pneumatic Pressure Check

To perform the Pneumatic Pressure Check, proceed as follows:

- Connect the NIBP inflation hose to the monitor without a cuff connected to the hose.
- Obtain the male luer plug from the Calibration Kit and use it to close the open end of the NIBP inflation hose. Twist the plug one-quarter turn to make sure the connection is not leaking. This is essential for the test to be performed.
- Enter the Service menu as described on page 104 and follow the general instructions given there.
- Press the DOWN key until the parameter “Pneumatic Test” is selected.
- Press one of the NEXT keys to move the cursor and begin the test.

The monitor starts to pressurize the pneumatic system to approximately 180 mmHg and attempts to hold this pressure. The pressure value will be displayed for about 15 seconds.

At the completion of a successful Pressure Check, the message “Passed” will display and the monitor will beep two times.

If the monitor fails the Pressure Check, the message "Failed" will be displayed and the monitor will beep three times.

- Press one of the PREVIOUS keys to exit and return to the Pneumatic Test selection.

**Caution:** Should the monitor fail the Pneumatic Pressure Check the test should be repeated. Instead of blocking the inflation hose with the male Luer plug, a 500 ml volume should be connected. (Obtain a fixed volume 500 ml Pressure Cylinder from CAS, part number 01-02-0248)

- Remove the LUER plug from the end of the air hose and connect the 500 ml Pressure Cylinder instead.

Should the monitor fail the Pneumatic Pressure Check with the 500 ml Pressure Cylinder attached, it needs to be serviced.

**Calibration Check and Overpressure Test**

To perform a Calibration Check, proceed as follows:

- Connect the NIBP inflation hose to the monitor without a cuff connected to the hose.
- Enter the Service menu as described on page 104 and follow the general instructions given there.
- Press the DOWN key until the parameter “Manometer Mode” is selected.
- Press one of the NEXT keys to move the cursor and begin the test.

The pressure in the pneumatic system is being displayed. At this point it is 0 mmHg.

- Use the manometer inflation bulb to slowly inflate the system pausing for 30 seconds at the following points and verify calibration according to the following table:

| | |
|---|---|
| 0 mmHg: | ± 1 mmHg |
| 50 mmHg: | ± 4 mmHg |
| 100 mmHg: | ± 4 mmHg |
| 150 mmHg: | ± 4 mmHg |
| 200 mmHg: | ± 5 mmHg |

To test the Overpressure safety function, proceed as follows:

- Inflate the pressure slowly until 290 mmHg ±10 mmHg is reached. The monitor's NIBP display should stop updating and the message "OVER PRESS" will be displayed.

**Note:** The monitor needs to be serviced if it does not meet the test criteria.

### PIC Voltage

The monitor displays the DC voltage level being received by the PIC processor from the Power Supply Board.

To view the PIC Voltage level, proceed as follows:

- Connect the monitor to an external power source (AC Line power).
- Enter the Service menu as described on page 104 and follow the general instructions given there.

Verify the voltage being displayed is 12.00V ± 0.50V. If this is not the case, the monitor needs to be serviced.

## Other Checks

### Temperature Calibration Check

To perform a Temperature Calibration Check, a Temperature Test Jack is required. Refer to the CAS 750 Service Manual (CAS part number 21-02-0191) for more information.

**Caution:** Do not perform a Temperature Test while the monitor is being used on a patient.

To perform a Temperature Test, proceed as follows:

- Remove any installed temperature probe from the TEMP receptacle.
- Insert the Temperature Test Jack into the TEMP receptacle.
- The TEMP value should read 37.0 °C (±0.1 °C) or 98.6 °F (±0.1 °F).

**Note:** The monitor needs to be serviced if this check fails.

- Remove the Temperature Test Jack from the TEMP receptacle once the test is completed.

# Battery Replacement

## Removing the Battery

- Turn the monitor off and disconnect the power cord.
- Push down on the battery latch to unlock the battery door from the rear panel of the monitor.
- Carefully remove the battery pack from the rear panel of the monitor.




Figure 22: Removing the Battery Pack

### Inserting the Battery

- Align the Battery Pack guides with the bottom of the monitor.
- Slowly close the battery door to ensure the connector in the monitor and the connector on the battery pack mate together.
- Lock the battery door closed.

**Note:** When the battery pack is re-inserted, the monitor will automatically turn on.

**Warning:** Use of unapproved batteries will invalidate the product's warranty and may result in serious safety consequences for the patient and user.

**Warning:** Do not disassemble the battery pack or batteries. The batteries contain electrolytes, which can cause injury to eyes, skin and clothing.

**Note:** This product contains a rechargeable battery that is recyclable. Under various state and local laws, it may be illegal to dispose of this battery into the municipal waste stream. Check with your local authorities for instructions on recycling options in your area.

## Fuse Replacement

**Note:** The **CAS 750M** monitor models have a single fuse located inside the monitor. This fuse is not user replaceable.

The **CAS 750** monitor has a dual fuse power input receptacle. Both AC lines are fused.

**Caution:** For continued protection against fire hazard, replace only with identically rated fuses. Refer to paragraph "Power" on page 143.

A fuse may need to be replaced if the monitor is plugged into an electrical outlet but the Battery Power Visual Indicator is not illuminated green.

**Warning:** Before changing the fuse, unplug the power cord.

To replace fuses, proceed as follows:

- Turn the monitor off and disconnect the power cord.
- Press down on the locking tab, which holds the fuse holder in the power input receptacle. While holding down on the tab, pull the fuse holder out.
- Remove the fuses.
- Place new fuses directly into the fuse holder.
- Insert the fuse holder into the power input receptacle. There should be an audible click when it is secured.

## Software Versions

The monitor displays the current software revision of its operating system and that of the internal modules being used inside.

To view the software versions, proceed as follows:

- Enter the monitor's Service menu as described on page 104.
- Press and hold the HISTORY key for 2 seconds to display the software versions.




Figure 23: Software Versions

(1) The $SpO_2$ module is optional, in the case when it is not installed, no text message is shown. Installed $SpO_2$ technology may be Masimo $SpO_2$ or Nellcor $SpO_2$.

(2) The $CO_2$ module is optional, in the case when it is not installed, no text message is shown.

(3) The NIBP module is optional, in the case when it is not installed, no text message is shown.

(4) The Analog Output is optional, in the case when it is not connected, the "N/A" text message is shown. When it is connected, the numerical version shall be displayed.

- Press the RETURN key to exit and return to the Main screen.

**Note:** If no key is pressed for 60 seconds, the monitor will automatically exit the Versions menu and return to the Main screen.

## Storage

Refer to paragraph “Operating Environment” on page 142 for storage temperature information.

**Warning:** If it becomes necessary to store the monitor for longer than 6 months, remove the monitor’s battery pack and place the monitor in its original packing container if available.

**Note:** Batteries not charged and left in storage for more than six months could degrade and not recharge to full capacity.

# 12 Printer

**Note:** The Citizen Model CMP-10 Mobile Printer that is used with the CAS 750 Monitor has proprietary software installed in it for communication.

For more detailed information on the Citizen Model CMP-10 Mobile Printer, refer to the User's Manual that was supplied with the printer.

**Warning:** The CAS 750 Monitor has been tested with the Citizen CMP-10 Mobile printer to comply with IEC 60601-1-1 and is the only printer that is recommended to be used with the monitor. If another printer is to be used, the user must read the Warning "*Leakage Current Test* " on page 16 and follow the guidance given.

The Citizen CMP-10 Mobile Printer interfaces to the monitor via an Infrared Port or by using the direct connect RS232 cable (supplied with printer). The Infrared ports are located on the top of the printer and on the bottom front panel of the monitor.

**Caution:** For safe and proper usage of the external printer, please observe the following:

- Avoid places where fluid may enter the printer opening accidentally.
- Avoid places subject to high or low temperature extremes.
- Avoid dusty places and where corrosive gasses are generated.
- Never attempt to dismantle or repair the printer mechanism.
- Do not drop or bump the printer.

When handling the thermal paper:

- Store in a dark, cool and dry place.
- Do not place near organic solvents.
- Avoid contact with vinyl chloride films erasers or adhesive tapes for extended periods.
- Avoid exposure to high temperature, humidity, liquid, or sunlight.
- Always use specified thermal paper. Refer to "Other Accessories and Options" on page 150 for order information.

# Printer Controls and Indicators




Figure 24: Printer Controls and Indicators

1. Charge LED (red = charging, green = fully charged)
2. Power LED (green = power on, red/green blinking fast = end of paper, red/green blinking slow = print head overheated)
3. Infrared (IrDA) port
4. RS 232 serial interface
5. Power ON/OFF key
6. Line Feed key
7. Paper Cover Release Button (press down to open cover)
8. Paper Cutting Edge
9. Paper Cover

# Printer Operation

## Direct Connection

The monitor uses the DB9 connector, located on the rear panel of the monitor, to interface to the RS232 port on the Citizen CMP-10 Mobile Printer. Use the cable, which is supplied with the printer.

**Note:** When using the direct connect method; connect the printer's serial cable to the printer before turning the printer on.

## Infrared Connection

**Note:** When using the infrared port, disconnect the printer serial cable from the printer before turning the printer on.

Position the printer's infrared port window in-line with the infrared window of the monitor.

**Note:** When using the Infrared port, it is important to keep the two devices close together (less than three feet / one meter) and in-line to maintain proper communications. The monitor can perform a communication test to verify the Infrared connection. Refer to paragraph "IrDA Test" on page105.

**Note:** When using the IrDA port for printing, it is recommended to keep all items clear of the communications path between the monitor and the printer.

- To turn the printer on, press and hold the ON/OFF key for 1 second. The Power LED illuminates green.
- To turn the printer off, press and hold the ON/OFF key. The Power LED will illuminate red and change back to green. When the LED illuminates green, remove your finger from the key.

**Note:** If the monitor is turned on and the printer is already on, the CAS logo is printed automatically.

The CMP-10 Mobile Printer has an Auto Power Off feature. If the monitor and printer become separated and during 10 minutes no data has been sent and the Line Feed (LF) key has not been pressed, the printer will automatically shutoff.

The monitor periodically sends a "wake-up" message to the printer that will disable the Auto Power Off feature.

Sample printouts of both, Trend and Alarm Histories, are shown below.

```
750X Series Monitor
21-Mar-06 15:48

Patient:
_____________________________

Notes:
_____________________________

_____________________________

Trend Hist:
HR:MN  HR  RR %O2  PR CO2  RR
13:35  60  15  98  59  38  16
13:35 NIBP=120/80 (90) PR= 59
13:36  60  15  98  59  38  16
13:37  60  15  98  59  38  16
13:37 NIBP=112/78 (89) PR= 61
13:38  60  15  98  59  38  16
13:39* 60  15  98  59  38  16
13:39 NIBP=120/80 (90) PR= 58
13:40  60  15  98  59  38  16
```

```
750X Series Monitor
21-Mar-06 15:48

Patient:
_____________________________

Notes:
_____________________________

_____________________________

Alarm Hist:
HR:MN Alarm
13:39 Tachycardia:  200 BPM
```

Figure 25: History Sample Printouts

**Note:** An asterisk (*) appears in the History printout to indicate an Alarm has occurred during that one-minute interval.

To print waveforms:

- Press and hold the RETURN key for 2 seconds.

The waveforms that are being displayed will print based on the selections made for Print Traces in the “Parameters Menu”. Refer to paragraph “Parameters Menu” on page 40.

**Note:** If the $EtCO_2$ Print selection is “ON” and the Print Traces selection is ECG and the $EtCO_2$ connector is installed or removed, all traces shall be printed – overriding the ECG selection.

When the Print On Alarm setting in the Parameters menu is set to “ON”, the type of alarm is printed on the printout and the trace is annotated with an arrow marker to mark the point of the alarm.




Figure 26: Waveform Sample Printout

1. Header
2. Trace Names
3. Size Marker
4. Alarm Event Type
5. Numeric Information
6. Alarm Event Marker

## Charging the Printer Battery

The CMP-10 Mobile Printer is equipped with a rechargeable Lithium Ion (LiION) battery pack. When the printer detects a Low Battery condition, the message “Low Battery” is printed and an audio indicator sounds three times.

**Warning:** Charge the printer battery using the AC Adapter, Model TRC-09-1100-M from GROUP WEST, or equivalent, included with the printer.

- Plug the battery charger’s cord into the printer battery charger jack, located on the rear panel.
- Plug the charger into an AC wall outlet of the appropriate voltage.
- Verify the Charge LED is lit red.

Battery charge time is approximately 3 hours. Once the battery is fully charged, the Charge LED switches to green.

# Installing Paper

**Note:** A red line appears when the remaining supply of thermal paper becomes low.

- Switch the printer off.
- Press the Cover Open button to access the paper compartment. Remove any remaining paper before installing the new roll.
- Place the new paper roll as shown on the illustration and pull out enough paper to reach out over the control panel of the printer.
- Close the paper door.



Figure 27: Paper Installation

**Note:** Make sure that the paper is correctly placed. If it is tilted in one or another direction and does not come out straight from under the cover, open the door and reposition the roll again.

**Warning:** Do not touch the print head or paper cutter while replacing the printer paper.

# Removing the Battery Pack

**Warning:** Do not operate the printer or connect it to the monitor with the battery pack is removed.

**Warning:** Never change the battery pack while the battery charger is plugged in and/or the monitor is being operated.

- Switch the printer off.
- Disconnect the printer from the monitor and unplug the wall charger cord.
- Open the battery door by pressing in on the battery cover and pushing upward.



Figure 28: Opening the Battery Door

- Remove the battery cover.
- Remove the battery pack from the compartment and disconnect its connecting cable.

# Installing the new Battery Pack:




Figure 29: Installing the New Battery

- Connect the battery cable into the battery connector.
- Insert the battery and its connecting cable into the battery compartment.
- Replace the battery cover by sliding it in from the back of the printer and pushing down to lock it in place.

**Caution:** Be sure to place the battery cover firmly in its position after installing the new battery pack.

**Warning:** Do not disassemble the battery pack or batteries. The batteries contain electrolytes, which can cause injury to eyes, skin and clothing.

**Note:** Under various state and local laws, it may be illegal to dispose of this battery into the municipal waste stream. Check with your local authorities for instructions on recycling options in your area.

# 13 External Device Interfacing

## Nurse Call and RS232 Interface

**Note:** The monitor uses the same DB9 connector for the Nurse Call and the RS232 interface.




| Pin | Signal |
|---|---|
| 1 | Nurse Call (N.O.) |
| 2 | Serial Receive In |
| 3 | Serial Transmit Out |
| 4 | No Connection |
| 5 | Isolated Ground |
| 6 | Nurse Call (N.C.) |
| 7 | No Connection |
| 8 | No Connection |
| 9 | Nurse Call (common) |

Figure 30: DB9 Connector

### RS232 Interface

The monitor uses the DB9 connector to interface to the RS232 port on the Citizen CMP-10 Mobile Printer. Use the cable, which is supplied with the printer.

The RS232 interface may be used for data collection or service purposes. Refer to the documentation provided with each application.

**Note:** The CAS 750 Monitor is not equipped with interfaces to remote equipment or network(s) to duplicate alarms.

### Nurse Call Interface

The Nurse Call output is activated by any violation of a patient high or low alarm limit. The Nurse Call is not affected by the silencing/muting of the audio when pressing the Silence/Reset key.

The output is compatible with most Nurse Call Systems in that there is no polarity to the connection. The monitor provides two isolated relay switches, a closing and an opening contact. For normally open applications, the nurse call system needs to be connected to pins 1 and 9 of the DB9 connector. For normally closed applications, the nurse call system needs to be connected to pins 6 and 9.

Refer to paragraph “Nurse Call Interface” on page 143 for technical specifications.

**Warning:** Qualified service personnel only should install the Nurse Call Interface connection.

**Warning:** The interconnection to the Nurse Call system may increase the total leakage current. The user must read the Warning “*Leakage Current Test* “and follow the guidance given.

**Note:** Even though the Nurse Call Interface allows remote alarm indication, it does not replace appropriate bedside surveillance by trained clinicians.

# Mounting

There are several mounting devices (e.g. roll stand or swivel mount) available for the monitor. In all cases, the 4 screw threads at the back are used to attach the monitor to the mounting device.




Figure 31: Mounting Threads

1. Screw Thread (one of four)

**Note:** Follow the instructions provided with each specific mounting solution.

# 14 Appendix

## Symbols

The following is a summary of all symbols used on the monitor and accessories. Symbols may occur on the product or on its packaging.

### Front Panel Symbols

Power Indicator

POWER (On/Standby)

ALARMS (Silence/Reset)

NIBP (Start, Cancel, Menu)

RETURN (Main Screen, Freeze, Print)

PARAMETERS

LIMITS

HISTORY

AUDIO/VISUAL

UP

DOWN

## Screen Indicators

Heart Beat Indicator

Respiration Breath Indicator

 Battery Charge Level Indicator

Silence Indicator

Bar Graph Indicator

## Symbols near Accessory Connections

Patient connections are Type BF

Patient connections are Type CF

Communication Port RS232 Connector

ECG/RESP ECG/Respiration Input Connector

SpO2 Pulse Oximeter Probe Input Connector

MicroStream™ $CO_2$ Input Connector

$CO_2$ Scavenger Exhaust Port

MAXNIBP® NIBP Hose and Cuff Connector

TEMP Temperature Probe Input Connector

## Symbols on Monitor or Printer

CAUTION, read instructions before using.

Potential Equalization Post

This symbol appears here instead of on the unit.
The first two digits of the unit's serial number indicate the year of manufacture in the 21st century.

IPX1 Protection against ingress of water.

CE 0086 The CE Mark and Notified Body Registration Number signify the device has met all essential requirements of European Medical Device Directive 93/42/EEC.

Indicates this monitor is subject to the Waste Electrical and Electronic Equipment Directive in the European Union.

Alternating Current



Symbol for DC power polarity

Symbol for DC power

Recycling suggested

Class II equipment

Read instructions in Printer Manual

## Symbols on Packaging

Lot Number

Relative Humidity for Storage and Transport

Single Patient Use. Do Not Reuse

Storage and Transport Temperatures

Use By date (yyyy-mm)

# Warranty Policy

**MONITORS (CAS 750)**

All products are sold by CAS Medical Systems, Inc. under the warranties set forth in the following paragraphs. Such warranties are extended only with respect to the purchase of this product directly from CAS Medical Systems, Inc. or CAS's Authorized Distributors as new merchandise and are extended to the first buyer thereof, other than for resale.

The CAS 750 Monitor is warranted for a period of twenty-four (24) months. All products, excluding printers, if applicable, are warranted to be free from functional defects in materials and workmanship and to conform to the description of the product contained in the Operator/Service Manual, published specifications, and accompanying labels and/or inserts, provided that the same is properly operated under conditions of normal use in accordance with applicable safety and regulatory requirements, and that replacements and repairs are made in accordance with the instructions provided by CAS Medical Systems, Inc.

The same warranty conditions are made for a period of twelve (12) months with respect to printer and battery. A ninety (90) days warranty is provided for non-disposable accessories such as reusable $SpO_2$ sensors, reusable temperature probes and other accessories provided by CAS as part of the original purchase. CAS warrants disposable or single-patient-use products, including blood pressure cuffs, for out-of-box failure only. Reusable blood pressure cuffs are warranted for one year. Where the accessory is not a CAS Medical Systems, Inc. manufactured product, the manufacturer's own warranty applies. Warranty of accessories purchased separately from listed suppliers will be the responsibility of such listed suppliers. Damage to any part through misuse, neglect, or accident, or by affixing any accessories or attachments other than CAS, Statcorp, Masimo®, Nellcor®, Nonin®, Oridion®, YSI® and Welch Allyn® manufactured accessories or attachments, is not covered by this warranty.

The foregoing warranties shall not apply if the product has been configured, modified, adjusted or repaired other than by CAS Medical Systems, Inc. or by persons expressly authorized by CAS Medical Systems, Inc., or not in accordance with written instructions provided by CAS Medical Systems, Inc., or if the product has been subjected to misuse, negligence, or accident. This warranty is void if the printer is used with any paper other than that specified by CAS Medical Systems, Inc.

CAS reserves the right to perform warranty service operations in its own factory, at an authorized repair facility, or at the customers' site. CAS Medical Systems, Inc.'s sole and exclusive obligation and Buyer's sole and exclusive remedy under the above warranties, is limited to repairing or replacing, free of charge, a product which is reported in writing or via telephone to CAS Medical Systems, Inc., has a Return Material Authorization (RMA) number assigned and which is returned during normal business hours, transporting charges prepaid to:

CAS Medical Systems, Inc.
44 East Industrial Road
Branford, CT. 06405 USA
Telephone: +1 203 488 6056
Fax: +1 203 488 9438
E-mail: custsrv@casmed.com

CAS MEDICAL SYSTEMS, INC. SHALL NOT BE OTHERWISE LIABLE FOR ANY DAMAGES INCLUDING, BUT NOT LIMITED TO, INCIDENTAL DAMAGES, CONSEQUENTIAL DAMAGES OR SPECIAL DAMAGES.

THERE ARE NO EXPRESS OR IMPLIED WARRANTIES WHICH EXTEND BEYOND THE WARRANTIES HEREINABOVE SET FORTH. CAS MEDICAL SYSTEMS, INC. MAKES NO WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE WITH RESPECT TO THE PRODUCT OR PARTS THEREOF.

# Monitor Error Messages

**Clock Battery**

The monitor's internal clock battery is almost discharged.

- The monitor needs to be serviced.

**Dead Battery**

The battery is fully discharged.

- Recharge the battery for at least 5 hours.

**Low Battery**

The battery is almost discharged. At least 30 minutes of operation is available from when the message first appears.

- Recharge the battery as soon as possible.

**System Error**

A failure has occurred on the Main Board.

- The monitor needs to be serviced.

**Power Failure**

Power was disconnected from the monitor.

- Press the ALARM key to clear the message.

OR

- Turn the monitor off and on again.

**SET CLOCK**

Time and date values are incorrect.

- Set the monitor's clock.
- If the condition persists after setting the monitors clock, then the monitor's internal clock battery needs to be replaced.

# Monitor Configurations

## CAS 750 Models with AC Power Supply

| | |
|---|---|
| CAS 750E-1 | ECG/Respiration, Temperature |
| CAS 750E-2 | ECG/Respiration, Temperature, MAXNIBP® |
| CAS 750E-2MS | ECG/Respiration, Temperature, Masimo $SpO_2$ |
| CAS 750E-2NL | ECG/Respiration, Temperature, Nellcor $SpO_2$ |
| CAS 750E-3MS | ECG/Respiration, Temperature, MAXNIBP®, Masimo $SpO_2$ |
| CAS 750E-3NL | ECG/Respiration, Temperature, MAXNIBP®, Nellcor $SpO_2$ |
| CAS 750E-3MSC | ECG/Respiration, Temperature, Oridion $CO_2$, Masimo $SpO_2$ |
| CAS 750E-3NLC | ECG/Respiration, Temperature, Oridion $CO_2$, Nellcor $SpO_2$ |
| CAS 750E-4MS | ECG/Respiration, Temperature, MAXNIBP®, Oridion $CO_2$, Masimo $SpO_2$ |
| CAS 750E-4NL | ECG/Respiration, Temperature, MAXNIBP®, Oridion $CO_2$, Nellcor $SpO_2$ |
| CAS 750C-2MS | Oridion $CO_2$, Masimo $SpO_2$ |
| CAS 750C-2NL | Oridion $CO_2$, Nellcor $SpO_2$ |
| CAS 750C-3MS | Oridion $CO_2$, Masimo $SpO_2$, MAXNIBP® |
| CAS 750C-3NL | Oridion $CO_2$, Nellcor $SpO_2$, MAXNIBP® |
| CAS 750S-2MS | Masimo $SpO_2$, MAXNIBP® |
| CAS 750S-2NL | Nellcor $SpO_2$, MAXNIBP® |

Table 4: CAS 750 Models with AC Power

## CAS 750 Models with 12 VDC Power Input

| | |
|---|---|
| CAS 750EM-1 | ECG/Respiration, Temperature |
| CAS 750EM-2 | ECG/Respiration, Temperature, MAXNIBP® |
| CAS 750EM-2MS | ECG/Respiration, Temperature, Masimo $SpO_2$ |
| CAS 750EM-2NL | ECG/Respiration, Temperature, Nellcor $SpO_2$ |
| CAS 750EM-3MS | ECG/Respiration, Temperature, MAXNIBP®, Masimo $SpO_2$ |
| CAS 750EM-3NL | ECG/Respiration, Temperature, MAXNIBP®, Nellcor $SpO_2$ |
| CAS 750EM-3MSC | ECG/Respiration, Temperature, Oridion $CO_2$, Masimo $SpO_2$ |
| CAS 750EM-3NLC | ECG/Respiration, Temperature, Oridion $CO_2$, Nellcor $SpO_2$ |
| CAS 750EM-4MS | ECG/Respiration, Temperature, MAXNIBP®, Oridion $CO_2$, Masimo $SpO_2$ |
| CAS 750EM-4NL | ECG/Respiration, Temperature, MAXNIBP®, Oridion $CO_2$, Nellcor $SpO_2$ |
| CAS 750CM-2MS | Oridion $CO_2$, Masimo $SpO_2$ |
| CAS 750CM-2NL | Oridion $CO_2$, Nellcor $SpO_2$ |
| CAS 750CM-3MS | Oridion $CO_2$, Masimo $SpO_2$, MAXNIBP® |
| CAS 750CM-3NL | Oridion $CO_2$, Nellcor $SpO_2$, MAXNIBP® |
| CAS 750SM-2MS | Masimo $SpO_2$, MAXNIBP® |
| CAS 750SM-2NL | Nellcor $SpO_2$, MAXNIBP® |

Table 5: CAS 750 Models with DC Power

# Monitor Configuration Record

**Monitor Model**

CAS 750 …….

CAS 750M…….

Serial Number: ……

**Installed Option Parameters**

- ECG, Respiration and Temperature
- NIBP
- $SpO_2$
- $CO_2$

**Main Screen Defaults**

Channel 1 :……

Channel 2 :……

Channel 3 :……

**Alarm Limit Defaults**

| Parameters | Adult | | Neonate | |
|---|---|---|---|---|
| | High | Low | High | Low |
| Heart Rate | | | | |
| %$SpO_2$ | | | | |
| SatSeconds | | | | |
| Pulse Rate | | | | |
| $EtCO_2$ | | | | |
| $FiCO_2$ | | | | |
| Resp. Rate | | | | |
| No Resp. | | | | |
| Systolic | | | | |
| Diastolic | | | | |

Date: _____ Unit: _____

# Specifications

## ECG

| | |
|---|---|
| Input: | 3 or 5 leads |
| Leads: | I, II, III, aVL, aVR, aVF, V |
| Frequency Response: | 0.5 to 40 Hz |
| Common Mode Rejection: | > 90 dB |
| Gain: | Automatic or selectable (X.5, X1, X2, X4, X8 or X16) |
| Pacer Detection: | Yes, meets AAMI Standards |
| Pacer Rejection: | Yes, meets AAMI Standards |
| Heart Rate Range: | Adult: 0, 15 - 300 BPM<br>Neonate: 0, 15 - 350 BPM |
| Accuracy: | ± 3 BPM or ± 3%, whichever is greater |
| Resolution: | 1 BPM |
| Update: | 1 second |
| Sensitivity Adjustment: | Automatic |
| Minimum Signal: | Neonate mode: 0.20 mV peak-to-peak<br>Pediatric/Adult mode: 0.20 mV peak-to-peak |
| Waveform Display: | QRS waveform, fixed for trace 1 (750E) |
| Display Sweep Speeds: | 12.5, 25, 50 mm/Sec |
| Patient Alarms: | |
| High HR: | Adjustable |
| Low HR: | Adjustable |
| Asystole: | 4 seconds no QRS |
| Equipment Alarms: | |
| Loose Lead: | Threshold at 2.0 to 2.5 K ohms |
| Alarm Delay Time: | 5 seconds |
| QRS Tone Volume: | Adjustable in 5 steps and Off |

## Temperature

| | |
|---|---|
| Input: | YSI® 400 compatible |
| Range: | 28 to 43 °C (82.4 to 109.4 °F) |
| Accuracy: | ± 0.1 °C (±0.2°F) |
| Resolution: | 0.1 °C or 0.1 °F |

## Respiration

| | |
|---|---|
| Method: | Impedance Pneumography |
| Sensing Electrodes: | LA-RA |
| Sense Frequency: | 63 kHz |
| Base Impedance Range: | Adult mode: 0 to 3.5 K ohms<br>Neonate mode: 0 to 2.5 K ohms |
| Bandwidth: | 0.05 to 2.5 Hz |
| Sensitivity Adjustment: | Automatic |
| Trigger Level: | Minimum trigger at 0.20 ohm. |
| Respiration Rate: | |
| Range in Neonate mode: | 0 (No RESP), 12-150 BrPM |
| Range in Adult mode: | 0 (No RESP), 6-150 BrPM |
| Accuracy: | 0-122 BrPM, ± 1 BrPM<br>123-150 BrPM, ± 2 BrPM |
| Resolution: | 1 BrPM |
| Update: | 1 second |
| Waveform Display: | Selectable for trace 2 or 3 |
| Display Sweep Speeds: | 3, 6.25, 12.5 mm/Sec |
| Patient Alarms: | |
| High RR: | Adjustable, 5 seconds delay |
| Low RR: | Adjustable, 5 seconds delay |
| No Respiration: | Adjustable from 10 Sec. to 60 Sec. |

## Oximetry

**Masimo SET®**

Type: Functional Oxygen Saturation

Wavelengths:

Red: 660 Nanometers

Infrared: 905 Nanometers

Radiant Power: 0.79mW maximum at 50 mA, pulsed

Measurement Range:

$\%SpO_2$: 0-100%

Pulse Rate: 25-240 BPM

Accuracy:

$\%SpO_2$:

| Sensor | Accuracy |
|---|---|
| DC-195 | 70 - 100%, +/-2 digits (1 S.D.) |
| LNOP® Adt | |
| LNOP Adt Long | |
| LNOP DCI | |
| LNOP DCSC | |
| LNOP DC1P | |
| LNOP DC150 | |
| LNOP Ear | |
| LNOP Pdt | |

| Sensor | Accuracy |
|---|---|
| LNOP Neo | 70 - 100%, +/-3 digits (1 S.D.) |
| LNOP Neo PT | |
| LNOP NeoPT-L | |
| LNOP Neo-L | |
| LNOP Inf-L | |

Pulse Rate: ± 3 BPM

Update of Numeric: every 1 second

Waveform Display: Plethysmogram, fixed for trace 1 (750C and 750S), selectable for trace 2 or 3

Sweep Speeds: 12.5, 25 or 50 mm/sec

Patient Alarms: High and low limits on $\%SpO_2$, adjustable. High and low limits on Pulse Rate, adjustable.

Alarm Delay: 0 or 10 seconds

**Nellcor® OxiMax®**

Type: Functional Oxygen Saturation

Wavelengths:

Red: 660 Nanometers

Infrared: 890 Nanometers

Power: Not exceeding 15 mW

Measurement Range:

$\%SpO_2$: 1-100%

Pulse Rate: 20-240 BPM

Accuracy:

$\%SpO_2$:

| Sensor | Accuracy |
|---|---|
| MAX-A<br>MAX-AL<br>MAX-N<br>MAX-P<br>MAX-I<br>MAX-FAST | 70 - 100%, +/-2 digits (1 S.D.) |

| Sensor | Accuracy |
|---|---|
| OxiCliq A<br>OxiCliq I<br>OxiCliq N<br>OxiCliq P | 70 - 100%, +/-2.5 digits (1 S.D.) |

| Sensor | Accuracy |
|---|---|
| DS-100A<br>D-YS<br>OXI-A/N<br>OXI-P/I | 70 - 100%, +/-3 digits (1 S.D.) |

| Sensor | Accuracy |
|---|---|
| D-YS<br>D-YSE<br>D-YSPD | 70 - 100%, +/-3.5 digits (1 S.D.) |

| Sensor | Accuracy |
|---|---|
| MAX-R | 80 - 100%, +/-3.5 digits (1 S.D.) |

Pulse: ± 3 digits

Updated of Numerics: every 1 second

Waveform Display: Plethysmogram, fixed for trace 1 (750C and 750S), selectable for trace 2 or 3

Sweep Speeds: 12.5, 25 or 50 mm/sec

Patient Alarms: High and low limits on $\%SpO_2$, adjustable.
High and low limits on Pulse Rate, adjustable.

Alarm Delay: $\%SpO_2$: SatSeconds™
Pulse: 0 or 10 seconds

## NIBP Measurement

| | |
|---|---|
| Technique: | Oscillometric (MAXNIBP Technology). |
| Patient Range: | Adult – Neonate |
| Self Test: | System self test is performed each time power is turned on. |
| Auto Zero: | Zero pressure reference is automatically established after every reading. |
| Inflation: | Initial inflation to 150 mmHg (Adult) or 85 mmHg (Neonatal) or user selectable. (Adaptive, 80, 100, 120, 140, 150, 160, 180, 200) - Adult; (Adaptive, 60, 80, 85, 100, 120) - Neonatal. Subsequent inflation to approximately 30 mmHg greater than previous Systolic pressure. |
| Deflation: | Automatically after measurement.<br>If cuff pressure exceeds 290 mmHg (Adult)<br>If cuff pressure exceeds 145 mmHg (Neonate)<br>If measurement time exceeds 120 seconds (Adult), 90 seconds (Neonate)<br>If safety timer detects microprocessor failure |
| Max Measurement Time: | Limited to 120 seconds (Adult), 90 seconds (Neonate) |
| Overpressure: | If cuff pressure exceeds 290 mmHg (Adult), 145 mmHg (Neonate) |

| Measurement Range: | Adult | Neonate |
|---|---|---|
| Systolic: | 30-255 mmHg | 30-135 mmHg |
| Diastolic: | 15-220 mmHg | 15-110 mmHg |
| MAP: | 20-235 mmHg | 20-125 mmHg |
| Pulse Rate Range: | 30-240 BPM | 40-240 BPM |

| | |
|---|---|
| Accuracy: | |
| Blood Pressure: | ± 5 mmHg, Standard deviation no greater than 8 mmHg. |
| Pulse Rate: | ± 2% or ± 2 BPM, whichever is greater. |
| Patient Alarms: | High and low limits for Systolic Pressure, adjustable.<br>High and low limits for Diastolic Pressure, adjustable. |

Accuracy complies with that given in American National Standard for Electronic or Automated Sphygmomanometers, ANSI/AAMI SP10, 2002. Adult blood pressure measurements determined with this device are equivalent to those obtained by an auscultatory blood pressure measurement device and neonatal ones are equivalent to those obtained by an intra-arterial blood pressure device, within the limits prescribed by the American National Standard for Electronic or Automated Sphygmomanometers. The 4th Korotkoff sound was used to determine Diastolic pressure. Study findings are available.

## Capnography

| | |
|---|---|
| Method: | Sidestream Capnography (MicroStream®) |
| Patient Range: | Adult - Neonate |
| Units: | mmHg, % or kPa |
| Sample Rate: | 50 ml/min ± 7.5 ml/min |
| System Response Time: | 2.9 seconds typical, (includes delay and rise time) |
| Rise Time: | Adult and Neonate: 190 msec max. |
| Delay Time: | 2.7 seconds (typical) |
| Frequency Response: | $EtCO_2$ accuracy is maintained up to 80 breaths/min. (For maintaining accuracy for respiration rate over 60 bpm, use Neonate mode.) From 81 to 150 bpm accuracy is ±12%, if the $EtCO_2$ is higher than 18.8 mmHg in Neonate mode. |
| Ambient Pressure: | Compensated internally – automatic |
| Range: | 0-99 mmHg (0-13.2 kPa and 0-13.0 vol% at sea level) |
| Initialization Time: | 30 seconds (typical) |
| Accuracy: | |
| Before steady state: | 0 – 38 mmHg ± 4 mmHg<br>39 – 99 mmHg ± 12% of reading |
| Steady state: | 0 – 38 mmHg ± 2 mmHg<br>39 – 99 mmHg ± 5% of reading +0.08% for every 1 mmHg above 38 mmHg. |
| Resolution: | 1 mmHg, 0.1 % or 0.1 kPa |
| Respiration Rate: | 0 - 150 breaths/min |
| Waveform Display: | $CO_2$ waveform, selectable for trace 2 or 3 |
| Sweep Speeds: | 3, 6.25, 12.5 mm/sec |
| Patient Alarms: | High and low limits on $EtCO_2$, adjustable<br>High $FiCO_2$, adjustable |

## Patient Alarms

| | |
|---|---|
| Adjustable Alarms: | High and low alarms for Heart Rate, %$SpO_2$, Pulse Rate, $EtCO_2$, High $FiCO_2$, Respiration Rate, Systolic Pressure, Diastolic Pressure, No Respiration. |
| Fixed Alarms: | Asystole |
| Alarm Indicators: | Audible<br>Red Patient Alarm LED's<br>Yellow Equipment Alarm LED's<br>Text in Patient Message Window |

| Patient Parameter | Neonatal Limit Range | | Adult Limit Range | |
|---|---|---|---|---|
| | Low | High | Low | High |
| Heart Rate: | 25 to 235 BPM | 25 to 235 BPM | 25 to 235 BPM | 25 to 235 BPM |
| %$SpO_2$: | 70 to 95 % | 80 to 99 % | 70 to 95 % | 80 to 99 % |
| SatSeconds: | 10 to 100 | N/A | 10 to 100 | N/A |
| $EtCO_2$: | 1 to 99 mmHg<br>.1 to 13.2 %<br>.1 to 13.2 kPa | 1 to 99 mmHg<br>.1 to 13.2 %<br>.1 to 13.2 kPa | 1 to 99 mmHg<br>.1 to 13.2 %<br>.1 to 13.2 kPa | 1 to 99 mmHg<br>.1 to 13.2 %<br>.1 to 13.2 kPa |
| $FiCO_2$: | N/A | 3 to 30 mmHg,OFF<br>.4 to 4.0 %,OFF<br>.4 to 4.0 kPa,OFF | N/A | 3 to 30 mmHg,OFF<br>.4 to 4.0 %,OFF<br>.4 to 4.0 kPa,OFF |
| Respiration Rate: | 5 to 145 BrPM | 5 to 145 BrPM | 5 to 145 BrPM | 5 to 145 BrPM |
| No Respiration: | 5 to 30 Sec | N/A | 5 to 30 Sec | N/A |
| SYS: | 35 to 130 mmHg | 35 to 130 mmHg | 35 to 250 mmHg | 35 to 250 mmHg |
| DIA: | 20 to 105 mmHg | 20 to 105 mmHg | 20 to 215 mmHg | 20 to 215 mmHg |

Each alarm limit may also be selected "OFF" individually or as a whole.
Low Limits cannot be set above the associated High Limit.
High Limits cannot be set lower than the associated Low Limit.

## Operating Modes

| | |
|---|---|
| Patient Modes: | Neonate or Adult |
| ECG: | Continuous Monitoring |
| Respiration: | Impedance Pneumography |
| NIBP: | Manual, STAT or Automatic (at preset intervals). |
| %$SpO_2$: | Continuous Monitoring |
| $CO_2$: | Continuous Monitoring |
| Temperature: | Continuous Monitoring |
| History: | Trend: Review of previous measurements |
| | Alarm: Review of previous alarms |
| Other Modes: | Stand-by mode and Auto Dim (based on model) |

## Display

| | |
|---|---|
| Display: | LCD display of measurement results, instructions, troubleshooting messages, waveforms and signal strength bar. |
| Numerics: | HR, PR, %$SpO_2$, NIBP (Systolic, Diastolic and MAP or PR), Et$CO_2$ (mmHg, % or kPa), RR, Temp (in °C or °F). |
| Waveforms: | Up to 3 traces, ECG, $SpO_2$, Impedance Respiration, $CO_2$. |
| Sweep Speeds: | 12.5, 25 or 50 mm/s for ECG and $SpO_2$.<br>3.0, 6.25 or 12.5 mm/s for Resp. and $CO_2$. |
| Trend History: | 480 1-minute entries of all parameters. |
| Alarm History: | 25 most recent alarms |

## Physical Dimensions and Weight

(Fully configured unit with battery)

| | |
|---|---|
| H x W x D: | 170 mm x 215 mm x 102 mm<br>(6.75 in x 8.5 in x 4.0 in) |
| Weight: | 2.0 kg (4.4 lbs) |

## Operating Environment

| | |
|---|---|
| Operating Temperature: | 0 to 50°C (32°F to 122°F) |
| Humidity: | 15 to 95% RH, non-condensing |
| Altitude: | 10,000 to –1,250 ft (700 to 1050 hPa) |

Monitors may not meet performance specifications if stored or used outside temperature and humidity ranges. When moving the monitor from a storage location, wait at least one-hour prior to use to allow the monitor to adjust to room temperature.

## Storage/Transport Environment

| | |
|---|---|
| Storage/Transport Temperature: | -20 to 60°C (-4°F to 149°F) |
| Humidity: | 15 to 95% RH, non-condensing |
| Altitude: | 10,000 to –1,250 ft (700 to 1050 hPa) |

## Power

External Power:

AC: 100-240 VAC, 50/60 Hz, 0.5 A
Fuse Rating – T500mAL250V or T1.25AL250V (two provided)
Refer to monitor labeling for actual fuse rating

DC (optional): +12 VDC; 7W
Fuse Rating – Littelfuse Type 154, Slo-Blo, 5.0A, 125VAC or approved equivalent (one provided)

Chassis Leakage Current: 100 µA (maximum)

Battery: NiMH battery pack (user removable).

Charge Time: 3-5 hours

Operating Time: 3 hrs (minimum)

## Serial Interface

| | |
|---|---|
| Interface Type: | Bi-directional Serial Communication |
| Speed: | 9600 for Printer<br>115200 for CAS Serial Protocol |
| Signal Level: | RS232C |
| Data Length: | 8 bits |
| Start Bit: | 1 bit |
| Stop Bit: | 1 bit |
| Parity: | None |
| Flow Control: | None |

## Nurse Call Interface

| | |
|---|---|
| Relay contacts: | Floating, one opening and one closing contact. |
| Max. Switching Power: | 0.3 A at 120 VAC or 1.0 A at 30 VDC |
| Max. Delay: | < 0.5 seconds. |

## Standards

Units comply with the following requirements:

ANSI/AAMI SP-10: 2002
CE marking according to Directive 93/42/EEC
IEC 60601-1, EN 60601-1-2
IEC 60601-2-27, IEC 60601-2-30, IEC 60601-20-49
EN 864, EN 865
UL classified – UL 60601-1, CAN/CSA C22.2 No. 60601.1

All units covered by U.S. patent 4,796,184 and 5,022,403. Other patents pending.

# Certificates

## Electronic Emissions and Immunity

| The Model 750 Monitor is intended for use in the electromagnetic environment specified below. The customer or the user of the Model 750 Monitor should assure it is used in such an environment. | | |
|---|---|---|
| **Emissions Test** | **Compliance** | **Electromagnetic Environment** |
| RF emissions – CISPR 11 | Group 1 | The Model 750 Monitor uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. |
| RF emissions – CISPR 11 | Class B | The Model 750 Monitor is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes. |
| Harmonic emissions IEC 61000-3-2 | Class B | |
| Voltage fluctuations / flicker emissions | Complies | |

| **Immunity Test** | **IEC 60601 Test Level** | **Compliance Level** | **Electromagnetic Environment Guidance** |
|---|---|---|---|
| Electrostatic discharge (ESD) IEC 61000-4-2 | ±6 kV contact<br>±8 kV air | ±6 kV contact<br>±8 kV air | Floors should be wood concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%. |
| Electrical fast transient/burst IEC 61000-4-4 | ±2 kV for power supply lines<br>±1 kV for input/output lines | ±2 kV for power supply lines<br>±1 kV for input/output lines | Mains power quality should be that of a typical commercial or hospital environment. |
| Surge IEC 61000-4-5 | ±1 kV differential mode<br>±2 kV common mode | ±1 kV differential mode<br>±2 kV common mode | Mains power quality should be that of a typical commercial or hospital environment. |
| Voltage Dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11 | < 5% UT (>95% dip in UT) for 0.5 cycle.<br>40% UT (60% dip in UT) for 5 cycles.<br>70% UT (30% dip in UT) for 25 cycles.<br>< 5% UT (> 95% dip in UT) for 5 seconds. | < 5% UT (>95% dip in UT) for 0.5 cycle.<br>40% UT (60% dip in UT) for 5 cycles.<br>70% UT (30% dip in UT) for 25 cycles.<br>< 5% UT (> 95% dip in UT) for 5 seconds. | Mains power quality should be that of a typical commercial or hospital environment. If user of the model 750 requires continued operation during power mains interruptions, it is recommended that the model 750 be powered from an uninterruptible power supply or a battery. |
| Power frequency (50/60 Hz) magnetic field IEC 61000-4-8 | 3 A/m | 3 A/m | Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment. |
| NOTE: $U_T$ is the A.C. mains voltage prior to application of the test level. | | | |

## Guidance and Manufacturer's Declaration – Electromagnetic Immunity

The Model 750 Monitor is intended for use in the electromagnetic environment specified below. The customer or the user of the Model 750 Monitor should insure that it is used in such an environment.

| Immunity Test | IEC 60601 Test Level | Compliance Level | Electromagnetic Environment - Guidance |
|---|---|---|---|
| | | | Portable and mobile RF communications equipment should be used no closer to any part of the Model 750 Monitor, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.<br>**Recommended separation distance:** |
| Conducted RF<br>IEC 61000-4-6 | 3 Vrms<br>150 kHz to 80 MHz | 3 Vrms | $d = 1.2\sqrt{P}$ |
| Radiated RF<br>IEC 61000-4-3 | 3 V/m<br>80 MHz to 2.5 GHz | 3 V/m | $d = 1.2\sqrt{P}$ 80 MHz to 800 MHz<br>$d = 2.3\sqrt{P}$ 800 MHz to 2.5 GHz<br><br>Where $P$ is the maximum output power rating of the transmitter in watts according to the transmitter manufacturer and $d$ is the recommended separation distance in meters.<br><br>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey [a], should be less than the compliance level in each frequency range. [b]<br><br>Interference may occur in the vicinity of equipment marked with the following symbol:<br> |

NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is effected by absorption and reflection from structures, objects and people.

[a] Field strengths from fixed transmitters, such as base stations for radio (cellular / cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the Model 750 Monitor is used exceeds the applicable RF compliance level above, the Model 750 Monitor should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as re-orienting or relocating the Model 750 Monitor.

[b] Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3 V/m.

## Recommended Separation Distances Between Portable and Mobile RF Communications Equipment and the Model 750 Monitor

The Model 750 Monitor is intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the Model 750 Monitor can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the Model 750 Monitor as recommended below, according to the maximum output power of the communications equipment.

| Rated maximum output power of transmitter (Watts) | Separation distance according to frequency of transmitter (Meters) | | |
|---|---|---|---|
| | 150 kHz to 80 MHz<br>$d = 1.2\sqrt{P}$ | 80 MHz to 800 MHz<br>$d = 1.2\sqrt{P}$ | 800 MHz to 2.5 GHz<br>$d = 2.3\sqrt{P}$ |
| 0.01 | 0.12 | 0.12 | 0.23 |
| 0.1 | 0.38 | 0.38 | 0.73 |
| 1 | 1.2 | 1.2 | 2.3 |
| 10 | 3.8 | 3.8 | 7.3 |
| 100 | 12 | 12 | 23 |

For transmitters operating at a maximum output power not listed above, the recommended separation distance $d$ in meters can be estimated using the equation applicable to the frequency of the transmitter, where $P$ is the maximum output power rating of the transmitter in watts according to the transmitter manufacturer.
NOTE 1 At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies.
NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

## CE Marking Information

### Compliance

The CAS750 patient monitors bear the CE mark CE-0086 indicating conformity with the provisions of the Council Directive 93/42/EEC concerning medical devices and fulfills the essential requirements of Annex I of this directive.

### Exceptions

None.

# Accessories

Contact our Customer Service Department or see our website for the latest product information. Refer to Page 2, Contact Addresses for e-mail and phone number information.

## ECG Accessories

| Catalog No. | Type | Description |
|---|---|---|
| 01-02-0304 | | 3-lead ECG Trunk Cable, 3m, IEC/AAMI |
| 01-02-0305 | | 5-lead ECG Trunk Cable 3m, IEC/AAMI |
| 01-02-0308 | | 3-lead set with clips, 61cm/24", AAMI |
| 01-02-0309 | | 5-lead set with clips, 91cm/36", AAMI |
| 01-02-0310 | | 3-lead set with clips, 61cm/24", IEC |
| 01-02-0311 | | 5-lead set with clips, 91cm/36", IEC |
| 01-02-0380 | | 5-lead set with clips, 91 cm/36 inch, IEC |
| 01-02-0381 | | 5-lead set with clips, 91 cm/36 inch, AAMI |
| 01-05-0016 | | Pregelled Snap Electrodes, Adult (12 pgks of 3) |
| 01-05-0014 | | Pregelled Snap Electrodes, Infant (18 pkgs of 3) |
| 28-02-0445 | | Leadwire Detangler |

## Tuff-Cuff Blood Pressure Cuffs

Reusable Blood Pressure Cuffs

| Catalog No. | Type | Description |
|---|---|---|
| 01-01-0308 | CR 5216 | Large Adult, 16 cm x 42 cm |
| 01-01-0307 | CR5214 | Adult, 14 cm x 37.5 cm |
| 01-01-0306 | CR5212 | Small, Adult12 cm x 30 cm |
| 01-01-0249 | CR5209 | Child, 9 cm x 27 cm |
| 01-01-0248 | CR5207 | Small Child, 7 cm x 21 cm |
| 01-01-0258 | CR5206 | Infant, 6 cm x 18 cm |

## Safe-Cuff Blood Pressure Cuffs

Single-Patient-Use Blood Pressure Cuffs

| Catalog No. | Type | Description |
|---|---|---|
| 01-01-0245 | CD2060 | X-Large Adult, 20 cm x 50 cm |
| 01-01-0243 | CD1642 | Large Adult, 16 cm x 40 cm |
| 01-01-0241 | CD1437 | Adult, 14 cm x 35 cm |
| 01-01-0239 | CD1230 | Small Adult, 12 cm x 30 cm |
| 01-01-0237 | CD927 | Child, 9 cm x 22.5 cm |
| 01-01-0235 | CD618 | Infant, 6 cm x 15 cm |

## Inflation Hoses

| Catalog No. | Type | Description |
|---|---|---|
| 01-01-0495 | | Coiled NIBP Hose for Adult and Pediatric applications, ten feet long, Bayonet cuff |
| 01-02-0185 | | Straight NIBP Hose for Neonatal and Infant applications, six feet long. |
| 01-02-0131 | | Coiled NIBP Hose for Adult and Pediatric applications, ten feet long. |

## Masimo $SpO_2$ Accessories

| Catalog No. | Type | Description |
|---|---|---|
| 01-02-0178 | LNOP-DCI | Adult digit, reusable sensor |
| 01-02-0190 | LNOP-DCIP | Pediatric digit, reusable sensor |
| 01-02-0191 | LNOP-YI | Multisite reusable sensor |
| 01-02-0312 | LNOP DC-195 | Adult digit, reusable sensor |
| 01-02-0432 | LNOP TC-1 | Ear, reusable sensor |
| 01-02-0251 | LNOP-Adt | Adult, single patient adhesive sensor (box of 20) |
| 01-02-0252 | LNOP-Adt Long | Adult, single patient adhesive sensor, long cable (box of 20) |
| 01-02-0253 | LNOP-Pdt | Pediatric/Slender digit single patient adhesive sensor (box of 20) |
| 01-02-0254 | LNOP-Neo | Neonatal single patient adhesive sensor (box of 20) |
| 01-02-0255 | LNOP-NeoPt | Neonatal preterm single patient adhesive sensor (box of 20) |
| 01-02-0182 | PC 04 | Patient Cable, 4 feet |
| 01-02-0192 | PC 08 | Patient Cable, 8 feet |

## Nellcor $SpO_2$ Accessories

| Catalog No. | Type | Description |
|---|---|---|
| 01-02-0179 | DS-100A | Durasensor® DS-100A, adult finger-clip sensor |
| 01-02-0183 | DOC-10 | OxiMax Patient Cable, 10 feet |

## Temperature Accessories

| Catalog No. | Type | Description |
|---|---|---|
| 01-02-0326 | 427 | YSI Reusable Skin Temperature Probe, 3 meter |
| 01-02-0327 | 402 | YSI Reusable Esphogeal/Rectal Temperature Probe, 3 meter |
| 01-02-0339 | 4491 | YSI Esphogeal/Rectal Disposable Temperature Probe |
| 01-02-0340 | 4499 | YSI Disposable Skin Probe |
| 01-02-0341 | 4940 | YSI Interconnect Cable for Disposable Probes, 3 meter, Reusable |
| 18-02-0235 | | Temperature Adapter Cable, 0.5 meter |

**Note:** A Temperature Adapter Cable is required from the monitor to any YSI cable or probe.

## Capnography Accessories

| Catalog No. | Type | Description |
|---|---|---|
| 01-02-0295 | | $CO_2$ Calibration Gas |

## Other Accessories and Options

| Catalog No. | Description |
|---|---|
| 01-02-0395 | Replacement Power Cord, U.S.A. |
| 01-02-0386 | Replacement Power Cord, European |
| 01-02-0385 | Replacement Power Cord, Australian |
| 01-02-0384 | Replacement Power Cord, U.K. |
| 03-08-0386 | DC Power Cable |
| 03-08-0450 | Monitor Battery Pack |
| 09-01-0002 | Fuse (250V, 500mA, 5x20 mm, Slow Blow), 2 per monitor – per labeling |
| 09-01-0034 | Fuse (250V, 1.25A, 5x20 mm, Slow Blow), 2 per monitor – per labeling |
| 01-01-0047 | P9 Calibration Kit |
| 01-02-0248 | 500 ml Fixed Volume Cylinder |
| 01-02-0172 | Roll Stand with Basket, Domestic |
| 01-02-0297 | Roll Stand with Basket, International |
| 01-02-0173 | Swivel Mount Kit |
| 01-02-0243 | Universal Pole Mount |
| 01-02-0174 | Carry Bag |
| 01-02-0189 | Printer including Battery, RS232 Cable, Power Supply, Roll of Paper and Manual |
| 01-02-0181 | Printer Bracket Attachment for 01-02-0172 Roll Stand |
| 01-02-0301 | Printer Bracket Attachment for 01-02-0297 Roll Stand |
| 28-02-0077 | Printer Paper, one Roll |
| 01-02-0266 | Printer Power Supply Adapter Plug, European |
| 01-02-0267 | Printer Power Supply Adapter Plug, UK |
| 01-02-0268 | Printer Power Supply Adapter Plug, Australian |
| 21-02-0191 | CAS 750 Service Manual |